



**State
Police**

**Forensic
Investigation
Center**

ANDREW M. CUOMO
Governor

KEITH M. CORLETT
Superintendent

DR. RAY A. WICKENHEISER
Director

0.10% SIMULATOR SOLUTION RECORD

Report Date: October 24, 2019
Manufacturer: Guth Laboratories, Inc.
Lot Number: 19250
Expiration Date: 3/30/2020
Analysis Results: 122 mg Ethyl Alcohol / 100 mL of solution

I, Carol S Miller, tested simulator solution lot number 19250 by headspace gas chromatography and have determined that it contains the appropriate concentration of ethyl alcohol. This solution is hereby approved for use.

When this simulator solution is used with a properly operating breath testing instrument, it will provide a value of 0.10% within acceptable limits.

I, Carol S Miller, made the entries contained in this record at the time of each event recorded, or within a reasonable time thereafter.

Performed by:

Reviewed by:

Digitally signed under
ESRA by Carol S Miller on
2019.10.24 11:45:19 -04'00'

Digitally signed under
ESRA by Carrie A Kirkton on
2019.10.24 12:54:00 -04'00'

Carol S Miller
Forensic Scientist III
Breath Analysis Technical Supervisor/Operator #43419


Carrie A Kirkton
Supervisor of Forensic Services
Breath Analysis Technical Supervisor

CERTIFIED COPY (CPLR RULE 4518)

0.10% SIMULATOR SOLUTION RECORD

Lot Number 19250, dated October 24, 2019

I, Jennifer F Limoges, Associate Director of Forensic Science, Forensic Investigation Center, do hereby certify and authenticate, as provided by Rule 4518, subdivisions a and c, Civil Practice Law and Rules, that the document annexed hereto is a true and accurate representation of an electronic record of the New York State Police Forensic Investigation Center which has been delegated to my possession, custody and control by the Superintendent of the New York State Police. This record, once created, is stored in an electronic format that cannot, thereafter, be altered or modified. This record was made in the regular course of business of the New York State Police Forensic Investigation Center, it is the regular course of the Forensic Investigation Center business to make such record, and such record was made at the time of each event recorded in it or within a reasonable time thereafter.



Digitally signed under
ESRA by Jennifer F Limoges
on
2019.10.24 14:48:52 -04'00'


Jennifer F Limoges
Associate Director of Forensic Science
Forensic Investigation Center

CERTIFIED COPY (CPLR RULE 4518)

0.10% SIMULATOR SOLUTION RECORD

Lot Number 19250, dated October 24, 2019

I, Jennifer F Limoges, Associate Director of Forensic Science, Forensic Investigation Center, do hereby certify and authenticate, as provided by Rule 4518, subdivisions a and c, Civil Practice Law and Rules, that the document annexed hereto is a true and accurate representation of an electronic record of the New York State Police Forensic Investigation Center which has been delegated to my possession, custody and control by the Superintendent of the New York State Police. This record, once created, is stored in an electronic format that cannot, thereafter, be altered or modified. This record was made in the regular course of business of the New York State Police Forensic Investigation Center, it is the regular course of the Forensic Investigation Center business to make such record, and such record was made at the time of each event recorded in it or within a reasonable time thereafter.



Digitally signed under
ESRA by Jennifer F Limoges
on
2019.10.24 14:48:52 -04'00'

Jennifer F Limoges

Associate Director of Forensic Science
Forensic Investigation Center



**State
Police**

**Forensic
Investigation
Center**

ANDREW M. CUOMO
Governor

KEITH M. CORLETT
Superintendent

DR. RAY A. WICKENHEISER
Director

0.10% SIMULATOR SOLUTION RECORD

Report Date: October 24, 2019
Manufacturer: Guth Laboratories, Inc.
Lot Number: 19250
Expiration Date: 3/30/2020
Analysis Results: 122 mg Ethyl Alcohol / 100 mL of solution

I, Carol S Miller, tested simulator solution lot number 19250 by headspace gas chromatography and have determined that it contains the appropriate concentration of ethyl alcohol. This solution is hereby approved for use.

When this simulator solution is used with a properly operating breath testing instrument, it will provide a value of 0.10% within acceptable limits.

I, Carol S Miller, made the entries contained in this record at the time of each event recorded, or within a reasonable time thereafter.

Performed by:

Reviewed by:

Digitally signed under
ESRA by Carol S Miller on
2019.10.24 11:45:19 -04'00'

Digitally signed under
ESRA by Carrie A Kirkton on
2019.10.24 12:54:00 -04'00'

Carol S Miller
Forensic Scientist III
Breath Analysis Technical Supervisor/Operator #43419


Carrie A Kirkton
Supervisor of Forensic Services
Breath Analysis Technical Supervisor

CERTIFIED COPY (CPLR RULE 4518)

0.10% SIMULATOR SOLUTION RECORD

Lot Number 19180, dated September 3, 2019

I, Jennifer F Limoges, Associate Director of Forensic Science, Forensic Investigation Center, do hereby certify and authenticate, as provided by Rule 4518, subdivisions a and c, Civil Practice Law and Rules, that the document annexed hereto is a true and accurate representation of an electronic record of the New York State Police Forensic Investigation Center which has been delegated to my possession, custody and control by the Superintendent of the New York State Police. This record, once created, is stored in an electronic format that cannot, thereafter, be altered or modified. This record was made in the regular course of business of the New York State Police Forensic Investigation Center, it is the regular course of the Forensic Investigation Center business to make such record, and such record was made at the time of each event recorded in it or within a reasonable time thereafter.



Digitally signed under
ESRA by Jennifer F
Limoges on
2019.09.03 14:29:40 -04'00'

Jennifer F Limoges

Associate Director of Forensic Science
Forensic Investigation Center



**State
Police**

**Forensic
Investigation
Center**

ANDREW M. CUOMO
Governor

KEITH M. CORLETT
Superintendent

DR. RAY A. WICKENHEISER
Director

0.10% SIMULATOR SOLUTION RECORD

Report Date: September 3, 2019
Manufacturer: Guth Laboratories, Inc.
Lot Number: 19180
Expiration Date: 1/24/2020
Analysis Results: 122 mg Ethyl Alcohol / 100 mL of solution

I, Carrie A Kirkton, tested simulator solution lot number 19180 by headspace gas chromatography and have determined that it contains the appropriate concentration of ethyl alcohol. This solution is hereby approved for use.

When this simulator solution is used with a properly operating breath testing instrument, it will provide a value of 0.10% within acceptable limits.

I, Carrie A Kirkton, made the entries contained in this record at the time of each event recorded, or within a reasonable time thereafter.

Performed by:

Reviewed by:

Digitally signed under
ESRA by Carrie A Kirkton on
2019.09.03 10:40:09 -04'00'

Digitally signed under
ESRA by Jennifer F Limoges
on
2019.09.03 14:26:34 -04'00'

Carrie A Kirkton
Supervisor of Forensic Services
Breath Analysis Technical Supervisor

Jennifer F Limoges
Associate Director of Forensic Science
Breath Analysis Technical Supervisor



**State
Police**

**Forensic
Investigation
Center**

ANDREW M. CUOMO
Governor

KEITH M. CORLETT
Superintendent

DR. RAY A. WICKENHEISER
Director

0.10% SIMULATOR SOLUTION RECORD

Report Date: June 14, 2019
Manufacturer: Guth Laboratories, Inc.
Lot Number: 19130
Expiration Date: 11/28/2019
Analysis Results: 119 mg Ethyl Alcohol / 100 mL of solution

I, Carol S Miller, tested simulator solution lot number 19130 by headspace gas chromatography and have determined that it contains the appropriate concentration of ethyl alcohol. This solution is hereby approved for use.

When this simulator solution is used with a properly operating breath testing instrument, it will provide a value of 0.10% within acceptable limits.

I, Carol S Miller, made the entries contained in this record at the time of each event recorded, or within a reasonable time thereafter.

Performed by:

Reviewed by:

Digitally signed under
ESRA by Carol S Miller on
2019.06.17 09:41:29 -04'00'

Digitally signed under ESRA
by Carrie A Kirkton on
2019.06.17 14:36:53 -04'00'

Carol S Miller


Forensic Scientist III

Breath Analysis Technical Supervisor/Operator #43419

Carrie A Kirkton

Supervisor of Forensic Services

Breath Analysis Technical Supervisor




CERTIFIED COPY (CPLR RULE 4518)

0.10% SIMULATOR SOLUTION RECORD

Lot Number 19130, dated June 14, 2019

I, Jennifer F Limoges, Associate Director of Forensic Science, Forensic Investigation Center, do hereby certify and authenticate, as provided by Rule 4518, subdivisions a and c, Civil Practice Law and Rules, that the document annexed hereto is a true and accurate representation of an electronic record of the New York State Police Forensic Investigation Center which has been delegated to my possession, custody and control by the Superintendent of the New York State Police. This record, once created, is stored in an electronic format that cannot, thereafter, be altered or modified. This record was made in the regular course of business of the New York State Police Forensic Investigation Center, it is the regular course of the Forensic Investigation Center business to make such record, and such record was made at the time of each event recorded in it or within a reasonable time thereafter.



Digitally signed under
ESRA by Jennifer F
Limoges on
2019.06.21 06:54:46
-04'00'

Jennifer F Limoges

Associate Director of Forensic Science
Forensic Investigation Center

