



**Division of Criminal
Justice Services**

6/25/2020

**CERTIFICATION PURSUANT TO CPLR 4518 OF RECORDS
MAINTAINED IN THE REGULAR COURSE OF BUSINESS**

I, Joshua Vinehout, Supervisor of Public Safety Programs, Office of Public Safety, New York State Division of Criminal Justice Services, 80 South Swan Street, Albany, New York 12210, having been duly designated and authorized by the Commissioner of the Division of Criminal Justice Services, do hereby certify and authenticate, as provided by subdivision c of Rule 4518 of the Civil Practice Law and Rules, that the document annexed hereto is an exact copy of an electronic record of the Office of Public Safety which is in my possession, custody and control. This record, once created, is stored in an electronic format that cannot thereafter be altered or modified.

I further certify that the original electronic record of inspection/maintenance/calibration of NAT'L PATENT ANALYTICAL SYS, DATAMASTER DMT breath test instrument, serial number 116306, performed on June 25, 2020 by Office of Public Safety employee Joseph Iavarone, of which the annexed is an exact copy, was made in the regular course of business of the New York State Division of Criminal Justice Services, Office of Public Safety, that such record was made at the time such inspection/ maintenance/calibration was performed or within a reasonable time thereafter, and further that it is the regular course of the Office of Public Safety's business to make such records at the time such inspection/maintenance /calibration is performed, or within a reasonable time thereafter, and to provide such instrument's records to the agency that requested them.

Digitally signed under
ESRA by Joshua Vinehout
on 6/26/2020 12:17 PM

**Joshua Vinehout
Supervisor of Public Safety Programs
Office of Public Safety**

An Equal Opportunity/Affirmative Action Agency

BREATH TEST INSTRUMENT RECORD OF INSPECTION / MAINTENANCE / CALIBRATION

Name of Submitting Agency: Ramapo Town Police Department

Manufacturer: NAT'L PATENT ANALYTICAL SYS

Model: DATAMASTER DMT

Date: June 25, 2020

Serial Number: 116306

Reference Solution Lot Number: 19281

Reference Solution Standard: 0.10 % at 34.0 degrees centigrade

Test No	Test Result	Simulator Temperature
1	0.100 %	33.99 degrees centigrade
2	0.100 %	33.99 degrees centigrade
3	0.100 %	33.99 degrees centigrade
4	0.100 %	33.99 degrees centigrade

Reference Solution Lot Number: 20020

Reference Solution Standard: 0.08 % at 34.0 degrees centigrade

Test No	Test Result	Simulator Temperature
1	0.081 %	33.99 degrees centigrade
2	0.080 %	33.99 degrees centigrade
3	0.081 %	33.99 degrees centigrade
4	0.081 %	33.99 degrees centigrade

Reference Solution Lot Number: 19210

Reference Solution Standard: 0.180 % at 34.0 degrees centigrade

Test No	Test Result	Simulator Temperature
1	0.183 %	33.99 degrees centigrade
2	0.183 %	34.00 degrees centigrade
3	0.183 %	33.99 degrees centigrade
4	0.182 %	34.00 degrees centigrade

I hereby certify that I have performed any necessary maintenance procedures and calibrated NAT'L PATENT ANALYTICAL SYS breath test instrument model DATAMASTER DMT, serial number 116306, and have determined that it is accurate and reliable for the determination of ethyl alcohol in the blood by analysis of the breath. The instrument's test results are recorded above and this instrument satisfies all limits and standards established by the New York State Commissioner of Health as enumerated in Part 59, Subchapter D of Chapter II, TITLE 10 (Health) of the Official Compilation of Codes, Rules and Regulations of the State of New York, (10 NYCRR Part 59) as amended.

I further certify that I made the entries in these records of inspection/maintenance/calibration at the time that such inspection/maintenance/calibration of the above identified breath test instrument was performed, or within a reasonable time thereafter.

Performed by:

Joseph Iavarone

Digitally signed under
ESRA by Joseph Iavarone
on 6/25/2020 12:14 PM

Joseph Iavarone
Highway Safety Equipment Technician
Breath Analysis Operator Permit # 44106

Reviewed by:

Frank Sgarlata


Digitally signed under
ESRA by Frank Sgarlata
on 6/26/2020 11:46 AM

Frank Sgarlata
Breath Analysis Technical Supervisor

**BREATH TEST INSTRUMENT
RECORD OF INSPECTION / MAINTENANCE / CALIBRATION**

Name of Submitting Agency: Ramapo Town Police Department
Manufacturer: NAT'L PATENT ANALYTICAL SYS
Model: DATAMASTER DMT
Serial Number: 116306
Date: June 25, 2020

- 1. General Appearance and condition of this device -----: OK
- 2. Check controls and indicators -----: OK
- 3. Check breath flow system -----: OK
- 4. Check printer operation -----: OK
- 5. System cycles through all modes -----: OK



6. Record any repairs made to this instrument and list services provided: INSPECTED INSTRUMENT, CHECKED SETUP VOLTAGES, PRINT AND REMOTE OPERATION. PERFORMED CALIBRATION/ADJUSTMENT PROCEDURE, RAN SUBJECT, TECHNICIAN AND DIAGNOSTIC TESTS, VERIFIED CALIBRATION.

Performed by:

Joseph Iavarone

Digitally signed under
ESRA by Joseph Iavarone
on 6/25/2020 12:14 PM

Joseph Iavarone
Highway Safety Equipment Technician
Breath Analysis Operator Permit # 44106

Reviewed by:

Frank Sgarlata

Digitally signed under
ESRA by Frank Sgarlata
on 6/26/2020 11:46 AM

Frank Sgarlata
Breath Analysis Technical Supervisor

TECHNICIAN TEST

RAMAPO TOWN
 POLICE DEPT
 DATAMASTER dmt: 116306


Date: 06/25/2020

OPERATOR NAME:
 JOSEPH A IAVARONE
 Permit #: 44106
 Operator Agency/Dept: RAMAPO TOWN PD
 Reference Standard #: 19281

Nominal = 0.100 g/210L

BLANK TEST	0.000	10:00
INTERNAL STANDARD	VERIFIED	10:00
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.100	10:01
BLANK TEST	0.000	10:02
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.100	10:02
BLANK TEST	0.000	10:03
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.100	10:03
BLANK TEST	0.000	10:04
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.100	10:04
BLANK TEST	0.000	10:05

Average = 0.1000
 Std Dev = 0.0000



TECHNICIAN TEST

RAMAPO TOWN
POLICE DEPT
DATAMASTER dmt: 116306


Date: 06/25/2020

OPERATOR NAME:
JOSEPH A LAVARONE
Permit #: 44106
Operator Agency/Dept: RAMAPO TOWN PD
Reference Standard #: 20020

Nominal = 0.080 g/210L

BLANK TEST	0.000	10:46
INTERNAL STANDARD	VERIFIED	10:46
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.081	10:46
BLANK TEST	0.000	10:47
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.080	10:48
BLANK TEST	0.000	10:48
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.081	10:49
BLANK TEST	0.000	10:50
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.081	10:50
BLANK TEST	0.000	10:51

Average = 0.0808
Std Dev = 0.0005



TECHNICIAN TEST

RAMAPO TOWN
POLICE DEPT
DATAMASTER dmt: 116306





Date: 06/25/2020

OPERATOR NAME:
JOSEPH A IAVARONE
Permit #: 44106
Operator Agency/Dept: RAMAPO TOWN PD
Reference Standard #: 19210

Nominal = 0.180 g/210L

BLANK TEST	0.000	10:52
INTERNAL STANDARD	VERIFIED	10:52
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.183	10:53
BLANK TEST	0.000	10:54
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.183	10:54
BLANK TEST	0.000	10:55
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.183	10:55
BLANK TEST	0.000	10:56
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.182	10:57
BLANK TEST	0.000	10:57

Average = 0.1828
Std Dev = 0.0005



SERVICE AUTHORIZATION FORM
Highway Safety Technology Unit
80 South Swan Street
Albany, NY 12210
Ph: 518-485-7636 / 518-402-0689
Fax: 518-457-6869

THIS FORM IS USED TO AUTHORIZE THE DIVISION OF CRIMINAL JUSTICE SERVICES TO PERFORM SERVICE ON THE INSTRUMENT IDENTIFIED. THIS AUTHORIZATION MUST ACCOMPANY EACH PIECE OF EQUIPMENT WHEN IT IS PRESENTED TO THE EQUIPMENT REPAIR CENTER FOR SERVICE. **PLEASE SIGN AND DATE FORM.**

SECTION I - AGENCY INFORMATION

Name of Agency <i>Town of Ramapo Police Dept</i>		Today's Date <i>6/25/2020</i>	
Address <i>237 ROUTE 59</i>		City, State, ZIP <i>SUFFERN, NY 10901</i>	
Contact Person <i>Adrius F. Sanchez</i>		Contact Telephone <i>845-222-8455</i>	
Contact Email <i>SANCHEZAR@RAMAPOPD.ORG</i>		Best Way to Reach You During Appointment Period <i>cell phone - Above</i>	

SECTION II - INSTRUMENT INFORMATION

Instrument Manufacturer (Please Circle) NPAS DMT REPCO SIM Applied Concepts(Stalker) Decatur(Genesis) Kustom(Eagle/Falcon) MPH A/S FST Other:			
Model Number <i>Data Master</i>		Instrument Serial Number <i>116306</i>	
Antenna Serial Number (if applicable) Antenna 1-- / Antenna 2--		Type of maintenance (check appropriate) <input checked="" type="checkbox"/> Calibration Check <input type="checkbox"/> Repair	
Brief Description of Malfunction			
List all parts, cables and/or accessories being submitted. Tuning forks must accompany all radar units (if applicable). (Please Circle)			
DMT/SIM: <input checked="" type="checkbox"/> DMT <input type="checkbox"/> SIM <input checked="" type="checkbox"/> Tubes <input checked="" type="checkbox"/> Keyboard Other:		Radar: Control Unit Antenna(s) Antenna Cable(s) Power Cord Forks Other:	
I acknowledge I am signing this Service Authorization Form as an act and deed of said agency, organization, or municipality, and that I am duly authorized to sign same for the uses and purposes mentioned herein. PLEASE SIGN AND DATE.			
Signature <i>[Signature]</i>		Date <i>6/25/2020</i>	

DCJS USE ONLY

inspected instrument, checked setup voltages, print and remote operation, performed calibration, adjustment procedure, ran subject, technician and diagnostic tests, verified calibration.