

SUPERVISOR TEST




RAMAPO TOWN
POLICE DEPT
DATAMASTER dmt: 116306

Date: 01/29/2022

OPERATOR NAME:
ANDRES F SANCHEZ
Permit #: 42014
Operator Agency/Dept: RAMAPO TOWN PD
Reference Standard #: 21390

Nominal = 0.10 g/210L

BLANK TEST	0.00	22:56
INTERNAL STANDARD	VERIFIED	22:56
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	22:57
BLANK TEST	0.00	22:58
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	22:58
BLANK TEST	0.00	22:59
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	22:59
BLANK TEST	0.00	23:00



SUPERVISOR TEST

RAMAPO TOWN
POLICE DEPT
DATAMASTER dmt: 116306


Date: 01/24/2022

OPERATOR NAME:
ANDRES F SANCHEZ
Permit #: 42014
Operator Agency/Dept: RAMAPO TOWN PD
Reference Standard #: 21390

Nominal = 0.10 g/210L

BLANK TEST	0.00	12:29
INTERNAL STANDARD	VERIFIED	12:29
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	12:30
BLANK TEST	0.00	12:31
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	12:31
BLANK TEST	0.00	12:32
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	12:32
BLANK TEST	0.00	12:33

1/24/2022
471



SUPERVISOR TEST

RAMAPO TOWN
 POLICE DEPT
 DATAMASTER dmt: 116306




Date: 01/20/2022

OPERATOR NAME:
 ANDRES F SANCHEZ
 Permit #: 42014
 Operator Agency/Dept: RAMAPO TOWN PD
 Reference Standard #: 21390

Nominal = 0.10 g/210L

BLANK TEST	0.00	08:21
INTERNAL STANDARD	VERIFIED	08:22
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	08:22
BLANK TEST	0.00	08:23
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	08:23
BLANK TEST	0.00	08:24
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	08:24
BLANK TEST	0.00	08:25

1/20/2022
471



SUPERVISOR TEST

RAMAPO TOWN
POLICE DEPT
DATAMASTER dmt: 116306




Date: 01/15/2022

OPERATOR NAME:
ANDRES F SANCHEZ
Permit #: 42014
Operator Agency/Dept: RAMAPO TOWN PD
Reference Standard #: 21390

Nominal = 0.10 g/210L

BLANK TEST	0.00	18:58
INTERNAL STANDARD	VERIFIED	18:58
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	18:59
BLANK TEST	0.00	19:00
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	19:00
BLANK TEST	0.00	19:01
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	19:01
BLANK TEST	0.00	19:02

1/15/2022
② 474



SUPERVISOR TEST

RAMAPO TOWN
POLICE DEPT
DATAMASTER dmt: 116306




Date: 01/13/2022

OPERATOR NAME:
ANDRES F SANCHEZ
Permit #: 42014
Operator Agency/Dept: RAMAPO TOWN PD
Reference Standard #: 21390

Nominal = 0.10 g/210L

BLANK TEST	0.00	11:20
INTERNAL STANDARD	VERIFIED	11:20
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	11:20
BLANK TEST	0.00	11:21
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	11:21
BLANK TEST	0.00	11:22
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	11:22
BLANK TEST	0.00	11:23

1/13/2022
0471



SUPERVISOR TEST

RAMAPO TOWN
POLICE DEPT
DATAMASTER dmt: 116306


Date: 01/10/2022

OPERATOR NAME:
ANDRES F SANCHEZ
Permit #: 42014
Operator Agency/Dept: RAMAPO TOWN PD
Reference Standard #: 21390

Nominal = 0.10 g/210L

BLANK TEST	0.00	15:11
INTERNAL STANDARD	VERIFIED	15:11
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	15:12
BLANK TEST	0.00	15:13
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	15:13
BLANK TEST	0.00	15:14
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	15:14
BLANK TEST	0.00	15:15

1/10/22
0471



SUPERVISOR TEST

RAMAPO TOWN
POLICE DEPT
DATAMASTER dmt: 116306




Date: 01/07/2022

OPERATOR NAME:
ANDRES F SANCHEZ
Permit #: 42014
Operator Agency/Dept: RAMAPO TOWN PD
Reference Standard #: 21390

Nominal = 0.10 g/210L

BLANK TEST	0.00	09:58
INTERNAL STANDARD	VERIFIED	09:58
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	09:58
BLANK TEST	0.00	09:59
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	09:59
BLANK TEST	0.00	10:00
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	10:00
BLANK TEST	0.00	10:01

1/7/2022
②471



SOLUTION CHANGE

RAMAPO TOWN
POLICE DEPT
DATAMASTER dmt: 116306


Date: 01/03/2022

OPERATOR NAME:
ANDRES F SANCHEZ
Permit #: 42014
Operator Agency/Dept: RAMAPO TOWN PD
Reference Standard #: 21390

Nominal = 0.10 g/210L

BLANK TEST	0.00	13:31
INTERNAL STANDARD	VERIFIED	13:31
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	13:32
BLANK TEST	0.00	13:32
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	13:33
BLANK TEST	0.00	13:33
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	13:34
BLANK TEST	0.00	13:35

1/3/2022
@ 2/7/



SOLUTION CHANGE

RAMAPO TOWN
POLICE DEPT
DATAMASTER dmt: 116306

Date: 01/03/2022


OPERATOR NAME:
ANDRES F SANCHEZ
Permit #: 42014
Operator Agency/Dept: RAMAPO TOWN PD
Reference Standard #: 21390

Nominal = 0.10 g/210L

BLANK TEST	0.00	13:16
INTERNAL STANDARD	VERIFIED	13:16
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD INTERFERENCE		13:16

Blank Error

1/3/22
@471



SOLUTION CHANGE

RAMAPO TOWN
POLICE DEPT
DATAMASTER dmt: 116306

Date: 01/03/2022

OPERATOR NAME:
ANDRES F SANCHEZ
Permit #: 42014
Operator Agency/Dept: RAMAPO TOWN PD
Reference Standard #: 21390


Nominal = 0.10 g/210L

BLANK TEST	0.00	13:07
INTERNAL STANDARD	VERIFIED	13:08
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	INTERFERENCE	13:08

Interference Detected

1/3/2022

② 471



SOLUTION CHANGE

RAMAPO TOWN
POLICE DEPT
DATAMASTER dmt: 116306

Date: 01/03/2022

OPERATOR NAME:
ANDRES F SANCHEZ
Permit #: 42014
Operator Agency/Dept: RAMAPO TOWN PD
Reference Standard #: 21390

Nominal = 0.10 g/210L

BLANK TEST	0.00	12:57
INTERNAL STANDARD	VERIFIED	12:57
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	INTERFERENCE	12:57

Interference Detected

1/3/2022

①47

