

SOLUTION CHANGE

RAMAPO TOWN
POLICE DEPT
DATAMASTER dmt: 116306

Date: 01/03/2022


OPERATOR NAME:
ANDRES F SANCHEZ
Permit #: 42014
Operator Agency/Dept: RAMAPO TOWN PD
Reference Standard #: 21390

Nominal = 0.10 g/210L

BLANK TEST
INTERNAL STANDARD 0.00 12:57
REFERENCE STANDARD VERIFIED 12:57
REFERENCE STANDARD TEMP 34.00C
REFERENCE STANDARD INTERFERENCE 12:57

Interference Detected

1/3/2022
AH



SOLUTION CHANGE

RAMAPO TOWN
POLICE DEPT
DATAMASTER dmt: 116306
Date: 01/03/2022


OPERATOR NAME:
ANDRES F SANCHEZ
Permit #: 42014
Operator Agency/Dept: RAMAPO TOWN PD
Reference Standard #: 21390

Nominal = 0.10 g/210L

BLANK TEST
INTERNAL STANDARD 0.00 13:07
REFERENCE STANDARD VERIFIED 13:08
REFERENCE STANDARD TEMP 34.00C
REFERENCE STANDARD INTERFERENCE 13:08

Interference Detected

1/3/2022
0471



SOLUTION CHANGE

RAMAPO TOWN
POLICE DEPT
DATAMASTER dmt: 116306

Date: 01/03/2022

OPERATOR NAME:
ANDRES F SANCHEZ
Permit #: 42014
Operator Agency/Dept: RAMAPO TOWN PD
Reference Standard #: 21390


Nominal = 0.10 g/210L

BLANK TEST
INTERNAL STANDARD 0.00 13:16
REFERENCE STANDARD VERIFIED 13:16
REFERENCE STANDARD TEMP 34.00C
REFERENCE STANDARD INTERFERENCE 13:16

Blank Error

1/3/22
@471

Alcohol(g/210L)
Flow Rate(L/M)



SOLUTION CHANGE

RAMAPO TOWN
POLICE DEPT
DATAMASTER dmt: 116306



Date: 01/03/2022

OPERATOR NAME:
ANDRES F SANCHEZ
Permit #: 42014
Operator Agency/Dept: RAMAPO TOWN PD
Reference Standard #: 21390

Nominal = 0.10 g/210L

BLANK TEST	0.00	13:31
INTERNAL STANDARD	VERIFIED	13:31
REFERENCE STANDARD	34.00C	
REFERENCE STANDARD	0.10	13:32
BLANK TEST	0.00	13:32
REFERENCE STANDARD	34.00C	
REFERENCE STANDARD	0.10	13:33
BLANK TEST	0.00	13:33
REFERENCE STANDARD	34.01C	
REFERENCE STANDARD	0.10	13:34
BLANK TEST	0.00	13:35

1/3/2022
@ 2/71



SUPERVISOR TEST

RAMAPO TOWN
POLICE DEPT
DATAMASTER dmt: 116306




Date: 01/07/2022

OPERATOR NAME:
ANDRES F SANCHEZ
Permit #: 42014
Operator Agency/Dept: RAMAPO TOWN PD
Reference Standard #: 21390

Nominal = 0.10 g/210L

BLANK TEST	0.00	09:58
INTERNAL STANDARD	VERIFIED	09:58
REFERENCE STANDARD	34.00C	
REFERENCE STANDARD	0.10	09:58
BLANK TEST	0.00	09:59
REFERENCE STANDARD	34.01C	
REFERENCE STANDARD	0.10	09:59
BLANK TEST	0.00	10:00
REFERENCE STANDARD	34.01C	
REFERENCE STANDARD	0.10	10:00
BLANK TEST	0.00	10:01

1/7/2022
@474



SUPERVISOR TEST




RAMAPO TOWN
POLICE DEPT
DATAMASTER dmt: 116306
Date: 01/10/2022

OPERATOR NAME:
ANDRES F SANCHEZ
Permit #: 42014
Operator Agency/Dept: RAMAPO TOWN PD
Reference Standard #: 21390

Nominal = 0.10 g/210L

BLANK TEST		
INTERNAL STANDARD	0.00	15:11
REFERENCE STANDARD	VERIFIED	15:11
REFERENCE STANDARD	34.01c	
BLANK TEST	0.10	15:12
REFERENCE STANDARD	0.00	15:13
REFERENCE STANDARD	34.01c	
BLANK TEST	0.10	15:13
REFERENCE STANDARD	0.00	15:14
REFERENCE STANDARD	34.00c	
BLANK TEST	0.10	15:14
REFERENCE STANDARD	0.00	15:15

1/10/22
0474



SUPERVISOR TEST




RAMAPO TOWN
POLICE DEPT
DATAMASTER dmt: 116306
Date: 01/13/2022

OPERATOR NAME:
ANDRES F SANCHEZ
Permit #: 42014
Operator Agency/Dept: RAMAPO TOWN PD
Reference Standard #: 21390

Nominal = 0.10 g/210L

BLANK TEST	0.00	11:20
INTERNAL STANDARD	VERIFIED	11:20
REFERENCE STANDARD	34.01c	
REFERENCE STANDARD	0.10	11:20
BLANK TEST	0.00	11:21
REFERENCE STANDARD	34.00c	
REFERENCE STANDARD	0.10	11:21
BLANK TEST	0.00	11:22
REFERENCE STANDARD	34.00c	
REFERENCE STANDARD	0.10	11:22
BLANK TEST	0.00	11:23

1/13/2022
D474



SUPERVISOR TEST




RAMAPO TOWN
POLICE DEPT
DATAMASTER dmt: 116306
Date: 01/15/2022

OPERATOR NAME:
ANDRES F SANCHEZ
Permit #: 42014
Operator Agency/Dept: RAMAPO TOWN PD
Reference Standard #: 21390

Nominal = 0.10 g/210L

BLANK TEST	0.00	18:58
INTERNAL STANDARD	VERIFIED	18:58
REFERENCE STANDARD	34.01c	
REFERENCE STANDARD	0.10	18:59
BLANK TEST	0.00	19:00
REFERENCE STANDARD	34.00c	
REFERENCE STANDARD	0.10	19:00
BLANK TEST	0.00	19:01
REFERENCE STANDARD	34.00c	
REFERENCE STANDARD	0.10	19:01
BLANK TEST	0.00	19:02

1/15/2022
@ 474



SUPERVISOR TEST




RAMAPO TOWN
POLICE DEPT
DATAMASTER dmt: 116306
Date: 01/20/2022

OPERATOR NAME:
ANDRES F SANCHEZ
Permit #: 42014
Operator Agency/Dept: RAMAPO TOWN PD
Reference Standard #: 21390

Nominal = 0.10 g/210L

BLANK TEST	0.00	08:21
INTERNAL STANDARD	VERIFIED	08:22
REFERENCE STANDARD	34.00C	
REFERENCE STANDARD	0.10	08:22
BLANK TEST	0.00	08:23
REFERENCE STANDARD	34.01C	
REFERENCE STANDARD	0.10	08:23
BLANK TEST	0.00	08:24
REFERENCE STANDARD	34.00C	
REFERENCE STANDARD	0.10	08:24
BLANK TEST	0.00	08:25

1/20/2022
474



SUPERVISOR TEST




RAMAPO TOWN
POLICE DEPT
DATAMASTER dmt: 116306
Date: 01/24/2022

OPERATOR NAME:
ANDRES F SANCHEZ
Permit #: 42014
Operator Agency/Dept: RAMAPO TOWN PD
Reference Standard #: 21390

Nominal = 0.10 g/210L

BLANK TEST	0.00	12:29
INTERNAL STANDARD	VERIFIED	12:29
REFERENCE STANDARD	34.01c	
REFERENCE STANDARD	0.10	12:30
BLANK TEST	0.00	12:31
REFERENCE STANDARD	34.01c	
REFERENCE STANDARD	0.10	12:31
BLANK TEST	0.00	12:32
REFERENCE STANDARD	34.00c	
REFERENCE STANDARD	0.10	12:33
BLANK TEST	0.00	12:33

1/24/2022
471



SUPERVISOR TEST

RAMAPO TOWN
POLICE DEPT
DATAMASTER dmt: 116306




Date: 01/29/2022

OPERATOR NAME:
ANDRES F SANCHEZ
Permit #: 42014
Operator Agency/Dept: RAMAPO TOWN PD
Reference Standard #: 21390

Nominal = 0.10 g/210L

INTERNAL STANDARD	0.00	22:56
VERIFIED	34.00c	22:56
REFERENCE STANDARD	0.10	22:57
REFERENCE STANDARD	0.00	22:58
BLANK TEST	34.00c	22:58
REFERENCE STANDARD	0.10	22:58
REFERENCE STANDARD	0.00	22:59
BLANK TEST	34.00c	22:59
REFERENCE STANDARD	0.10	22:59
REFERENCE STANDARD	0.00	23:00
BLANK TEST		

1/29/22
AS 477



SOLUTION CHANGE




RAMAPO TOWN
POLICE DEPT
DATAMASTER dmt: 116306
Date: 02/04/2022

OPERATOR NAME:
ANDRES F SANCHEZ
Permit #: 42014
Operator Agency/Dept: RAMAPO TOWN PD
Reference Standard #: 21440

Nominal = 0.10 g/210L

BLANK TEST	0.00	12:24
INTERNAL STANDARD	VERIFIED	12:24
REFERENCE STANDARD	34.00C	
REFERENCE STANDARD	0.10	12:24
BLANK TEST	0.00	12:25
REFERENCE STANDARD	34.01C	
REFERENCE STANDARD	0.10	12:25
BLANK TEST	0.00	12:26
REFERENCE STANDARD	34.01C	
REFERENCE STANDARD	0.10	12:27
BLANK TEST	0.00	12:27

2/4/2022
Ph



SUPERVISOR TEST

RAMAPO TOWN
POLICE DEPT
DATAMASTER dmt: 116306




Date: 02/07/2022

OPERATOR NAME:
ANDRES F SANCHEZ
Permit #: 42014
Operator Agency/Dept: RAMAPO TOWN PD
Reference Standard #: 21440

Nominal = 0.10 g/210L

BLANK TEST	0.00	09:47
INTERNAL STANDARD	VERIFIED	09:47
REFERENCE STANDARD	34.00C	
REFERENCE STANDARD	0.10	09:47
BLANK TEST	0.00	09:48
REFERENCE STANDARD	34.01C	
REFERENCE STANDARD	0.10	09:48
BLANK TEST	0.00	09:49
REFERENCE STANDARD	34.01C	
REFERENCE STANDARD	0.10	09:49
BLANK TEST	0.00	09:50

2/7/2022
471



SUPERVISOR TEST

RAMAPO TOWN
POLICE DEPT
DATAMASTER dmt: 116306




Date: 02/10/2022

OPERATOR NAME:
ANDRES F SANCHEZ
Permit #: 42014
Operator Agency/Dept: RAMAPO TOWN PD
Reference Standard #: 21440

Nominal = 0.10 g/210L

BLANK TEST		
INTERNAL STANDARD	0.00	17:43
REFERENCE STANDARD	VERIFIED	17:43
REFERENCE STANDARD	34.00c	
BLANK TEST	0.10	17:43
REFERENCE STANDARD	0.00	17:44
REFERENCE STANDARD	34.01c	
BLANK TEST	0.10	17:44
REFERENCE STANDARD	0.00	17:45
REFERENCE STANDARD	34.00c	
BLANK TEST	0.10	17:46
REFERENCE STANDARD	0.00	17:46

2/10/22
@47



SUPERVISOR TEST




RAMAPO TOWN
 POLICE DEPT
 DATAMASTER dmt: 116306
 Date: 02/14/2022

OPERATOR NAME:
 ANDRES F SANCHEZ
 Permit #: 42014
 Operator Agency/Dept: RAMAPO TOWN PD
 Reference Standard #: 21440

Nominal = 0.10 g/210L

BLANK TEST	0.00	09:20
INTERNAL STANDARD	VERIFIED	09:20
REFERENCE STANDARD	34.00c	
REFERENCE STANDARD	0.10	09:20
BLANK TEST	0.00	09:21
REFERENCE STANDARD	34.01c	
REFERENCE STANDARD	0.10	09:21
BLANK TEST	0.00	09:22
REFERENCE STANDARD	34.00c	
REFERENCE STANDARD	0.10	09:23
BLANK TEST	0.00	09:23

2/14/22
 @477



SUPERVISOR TEST




RAMAPO TOWN
POLICE DEPT
DATAMASTER dmt: 116306
Date: 02/21/2022

OPERATOR NAME:
ANDRES F SANCHEZ
Permit #: 42014
Operator Agency/Dept: RAMAPO TOWN PD
Reference Standard #: 21440

Nominal = 0.10 g/210L

BLANK TEST	0.00	10:01
INTERNAL STANDARD	VERIFIED	10:01
REFERENCE STANDARD	34.01c	
REFERENCE STANDARD	0.10	10:01
BLANK TEST	0.00	10:02
REFERENCE STANDARD	34.01c	
REFERENCE STANDARD	0.10	10:02
BLANK TEST	0.00	10:03
REFERENCE STANDARD	34.00c	
REFERENCE STANDARD	0.10	10:03
BLANK TEST	0.00	10:04

2/24/22
@2471



SUPERVISOR TEST




RAMAPO TOWN
POLICE DEPT
DATAMASTER dmt: 116306
Date: 03/01/2022

OPERATOR NAME:
ANDRES F SANCHEZ
Permit #: 42014
Operator Agency/Dept: RAMAPO TOWN PD
Reference Standard #: 21440

Nominal = 0.10 g/210L

BLANK TEST	0.00	10:15
INTERNAL STANDARD	VERIFIED	10:15
REFERENCE STANDARD	34.01c	
REFERENCE STANDARD	0.10	10:15
BLANK TEST	0.00	10:16
REFERENCE STANDARD	34.00c	
REFERENCE STANDARD	0.10	10:17
REFERENCE STANDARD	0.00	10:17
BLANK TEST	34.01c	
REFERENCE STANDARD	0.10	10:18
REFERENCE STANDARD	0.00	10:19
BLANK TEST		

3/1/2022
D471



SOLUTION CHANGE

RAMAPO TOWN
POLICE DEPT
DATAMASTER dmt: 116306

Date: 03/01/2022


OPERATOR NAME:
ANDRES F SANCHEZ
Permit #: 42014
Operator Agency/Dept: RAMAPO TOWN PD
Reference Standard #: 21440

Nominal = 0.10 g/210L

BLANK TEST
INTERNAL STANDARD 0.00 10:39
VERIFIED 10:40
REFERENCE STANDARD TEMP 34.00C
REFERENCE STANDARD INTERFERENCE 10:40

Blank Error

3/1/22
D. Sanchez



SOLUTION CHANGE

RAMAPO TOWN
POLICE DEPT
DATAMASTER dmt: 116306




Date: 03/01/2022

OPERATOR NAME:
ANDRES F SANCHEZ
Permit #: 42014
Operator Agency/Dept: RAMAPO TOWN PD
Reference Standard #: 21440

Nominal = 0.10 g/210L

BLANK TEST	0.00	12:12
INTERNAL STANDARD	VERIFIED	12:12
REFERENCE STANDARD	34.00c	
REFERENCE STANDARD	0.10	12:13
BLANK TEST	0.00	12:13
REFERENCE STANDARD	34.01c	
REFERENCE STANDARD	0.10	12:14
BLANK TEST	0.00	12:15
REFERENCE STANDARD	34.00c	
REFERENCE STANDARD	0.10	12:15
BLANK TEST	0.00	12:16

3/1/2022
AS



TECHNICIAN TEST



RAMAPO TOWN
POLICE DEPT
DATAMASTER dmt: 116306
Date: 03/02/2022

Nominal = 0.100 g/210L

BLANK TEST	0.000	13:37
INTERNAL STANDARD	VERIFIED	13:37
REFERENCE STANDARD	34.00c	
REFERENCE STANDARD	0.103	13:37
REFERENCE STANDARD	0.000	13:38
BLANK TEST	34.01c	
REFERENCE STANDARD	0.103	13:38
REFERENCE STANDARD	0.000	13:39
BLANK TEST	34.00c	
REFERENCE STANDARD	0.103	13:39
REFERENCE STANDARD	0.000	13:40
BLANK TEST	34.01c	
REFERENCE STANDARD	0.104	13:40
REFERENCE STANDARD	0.000	13:41
BLANK TEST		

Average = 0.1032
Std Dev = 0.0005

3/2/2022
@ 471



SUPERVISOR TEST




RAMAPO TOWN
POLICE DEPT
DATAMASTER dmt: 116306
Date: 03/06/2022

OPERATOR NAME:
ANDRES F SANCHEZ
Permit #: 42014
Operator Agency/Dept: RAMAPO TOWN PD
Reference Standard #: 21440

Nominal = 0.10 g/210L

BLANK TEST	0.00	20:01
INTERNAL STANDARD	VERIFIED	20:01
REFERENCE STANDARD	34.00C	
REFERENCE STANDARD	0.10	20:01
BLANK TEST	0.00	20:02
REFERENCE STANDARD	34.00C	
REFERENCE STANDARD	0.10	20:03
BLANK TEST	0.00	20:03
REFERENCE STANDARD	34.00C	
REFERENCE STANDARD	0.10	20:04
BLANK TEST	0.00	20:05

3/6/2022
@474



SUPERVISOR TEST

RAMAPO TOWN
POLICE DEPT
DATAMASTER dmt: 116306




Date: 03/11/2022

OPERATOR NAME:
ANDRES F SANCHEZ
Permit #: 42014
Operator Agency/Dept: RAMAPO TOWN PD
Reference Standard #: 21440

Nominal = 0.10 g/210L

BLANK TEST	0.00	17:41
INTERNAL STANDARD	VERIFIED	17:41
REFERENCE STANDARD	34.00c	
REFERENCE STANDARD	0.10	17:42
REFERENCE STANDARD	0.00	17:43
BLANK TEST	34.01c	
REFERENCE STANDARD	0.10	17:43
REFERENCE STANDARD	0.00	17:44
BLANK TEST	34.00c	
REFERENCE STANDARD	0.10	17:44
REFERENCE STANDARD	0.00	17:45
BLANK TEST		

3/11/2022
@477



SUPERVISOR TEST




RAMAPO TOWN
 POLICE DEPT
 DATAMASTER dmt: 116306
 Date: 03/15/2022

OPERATOR NAME:
 ANDRES F SANCHEZ
 Permit #: 42014
 Operator Agency/Dept: RAMAPO TOWN PD
 Reference Standard #: 21440

Nominal = 0.10 g/210L

BLANK TEST	0.00	09:23
INTERNAL STANDARD	VERIFIED	09:23
REFERENCE STANDARD	34.00C	
REFERENCE STANDARD	0.10	09:23
BLANK TEST	0.00	09:24
REFERENCE STANDARD	34.00C	
REFERENCE STANDARD	0.10	09:24
BLANK TEST	0.00	09:25
REFERENCE STANDARD	34.00C	
REFERENCE STANDARD	0.10	09:25
BLANK TEST	0.00	09:26

3/15/2022
AS 977



SUPERVISOR TEST

RAMAPO TOWN
POLICE DEPT
DATAMASTER dmt: 116306




Date: 03/18/2022

OPERATOR NAME:
ANDRES F SANCHEZ
Permit #: 42014
Operator Agency/Dept: RAMAPO TOWN PD
Reference Standard #: 21440

Nominal = 0.10 g/210L

BLANK TEST	0.00	14:57
INTERNAL STANDARD	VERIFIED	14:57
REFERENCE STANDARD	34.00C	
REFERENCE STANDARD	0.10	14:57
BLANK TEST	0.00	14:58
REFERENCE STANDARD	34.00C	
REFERENCE STANDARD	0.10	14:58
BLANK TEST	0.00	14:59
REFERENCE STANDARD	34.00C	
REFERENCE STANDARD	0.10	14:59
BLANK TEST	0.00	15:00

3/18/2022
@971



SUPERVISOR TEST




RAMAPO TOWN
POLICE DEPT
DATAMASTER dmt: 116306
Date: 03/22/2022

OPERATOR NAME:
ANDRES F SANCHEZ
Permit #: 42014
Operator Agency/Dept: RAMAPO TOWN PD
Reference Standard #: 21440

Nominal = 0.10 g/210L

BLANK TEST	0.00	13:14
INTERNAL STANDARD	VERIFIED	13:14
REFERENCE STANDARD	34.01c	
REFERENCE STANDARD	0.10	13:14
BLANK TEST	0.00	13:15
REFERENCE STANDARD	34.01c	
REFERENCE STANDARD	0.10	13:16
BLANK TEST	0.00	13:16
REFERENCE STANDARD	34.00c	
REFERENCE STANDARD	0.10	13:17
BLANK TEST	0.00	13:18

3/22/22
Ⓟ974



SUPERVISOR TEST




RAMAPO TOWN
POLICE DEPT
DATA MASTER dmt: 116306
Date: 03/25/2022

OPERATOR NAME:
ANDRES F SANCHEZ
Permit #: 42014
Operator Agency/Dept: RAMAPO TOWN PD
Reference Standard #: 21440

Nominal = 0.10 g/210L

BLANK TEST	0.00	10:48
INTERNAL STANDARD	VERIFIED	10:48
REFERENCE STANDARD	34.01C	
REFERENCE STANDARD	0.10	10:49
BLANK TEST	0.00	10:49
REFERENCE STANDARD	34.00C	
REFERENCE STANDARD	0.10	10:50
BLANK TEST	0.00	10:51
REFERENCE STANDARD	34.00C	
REFERENCE STANDARD	0.10	10:51
BLANK TEST	0.00	10:52

3/25/2022
② 477



SUPERVISOR TEST

RAMAPO TOWN
POLICE DEPT
DATAMASTER dmt: 116306

Date: 03/29/2022


OPERATOR NAME:
ANDRES F SANCHEZ
Permit #: 42014
Operator Agency/Dept: RAMAPO TOWN PD
Reference Standard #: 21440

Nominal = 0.10 g/210L


INTERNAL STANDARD	0.00	03:27
REFERENCE STANDARD	34.00c	03:27
REFERENCE STANDARD	0.10	03:28
BLANK TEST	0.00	03:28
REFERENCE STANDARD	34.00c	
REFERENCE STANDARD	0.10	03:29
BLANK TEST	0.00	03:29
REFERENCE STANDARD	34.01c	
REFERENCE STANDARD	0.10	03:30
BLANK TEST	0.00	03:31

3/29/22
P47A


Alcohol (g/210L)
Flow Rate (L/M)



Alcohol (g/210L)
Flow Rate (L/M)



Alcohol (g/210L)
Flow Rate (L/M)






SUPERVISOR TEST

RAMAPO TOWN
 POLICE DEPT
 DATAMASTER dmt: 116306
 Date: 04/01/2022

OPERATOR NAME:
 ANDRES F SANCHEZ
 Permit #: 42014
 Operator Agency/Dept: RAMAPO TOWN PD
 Reference Standard #: 21440

Nominal = 0.10 g/210L

BLANK TEST	0.00	17:20
INTERNAL STANDARD	VERIFIED	17:21
REFERENCE STANDARD	34.00C	
REFERENCE STANDARD	0.10	17:21
BLANK TEST	0.00	17:22
REFERENCE STANDARD	34.01C	
REFERENCE STANDARD	0.10	17:22
BLANK TEST	0.00	17:23
REFERENCE STANDARD	34.01C	
REFERENCE STANDARD	0.10	17:23
BLANK TEST	0.00	17:24



SOLUTION CHANGE




RAMAPO TOWN
POLICE DEPT
DATAMASTER dmt: 116306
Date: 04/08/2022

OPERATOR NAME:
ANDRES F SANCHEZ
Permit #: 42014
Operator Agency/Dept: RAMAPO TOWN PD
Reference Standard #: 22030

Nominal = 0.10 g/210L

BLANK TEST	0.00	16:48
INTERNAL STANDARD	VERIFIED	16:49
REFERENCE STANDARD	34.00c	
REFERENCE STANDARD	0.10	16:49
BLANK TEST	0.00	16:50
REFERENCE STANDARD	34.00c	
REFERENCE STANDARD	0.10	16:50
BLANK TEST	0.00	16:51
REFERENCE STANDARD	34.01c	
REFERENCE STANDARD	0.10	16:51
BLANK TEST	0.00	16:52

4/8/22
AD 47A



SUPERVISOR TEST




RAMAPO TOWN
POLICE DEPT
DATAMASTER dmt: 116306
Date: 04/13/2022

OPERATOR NAME:
ANDRES F SANCHEZ
Permit #: 42014
Operator Agency/Dept: RAMAPO TOWN PD
Reference Standard #: 22030

Nominal = 0.10 g/210L

BLANK TEST	0.00	14:20
INTERNAL STANDARD	VERIFIED	14:20
REFERENCE STANDARD	34.01C	
BLANK TEST	0.10	14:20
REFERENCE STANDARD	0.00	14:21
REFERENCE STANDARD	34.01C	
BLANK TEST	0.10	14:21
REFERENCE STANDARD	0.00	14:22
REFERENCE STANDARD	34.01C	
BLANK TEST	0.10	14:22
REFERENCE STANDARD	0.00	14:23
REFERENCE STANDARD	34.01C	
BLANK TEST	0.00	14:23

2/13/22
@ 47A



SUPERVISOR TEST

RAMAPO TOWN
POLICE DEPT
DATAMASTER dmt: 116306
Date: 04/19/2022




OPERATOR NAME:
ANDRES F SANCHEZ
Permit #: 42014
Operator Agency/Dept: RAMAPO TOWN PD
Reference Standard #: 22030

Nominal = 0.10 g/210L

TEST TYPE	VERIFIED	TIME
INTERNAL STANDARD	0.00	18:28
REFERENCE STANDARD	34.01c	18:28
REFERENCE STANDARD	0.10	18:28
BLANK TEST	0.00	18:29
REFERENCE STANDARD	34.01c	18:29
REFERENCE STANDARD	0.10	18:29
BLANK TEST	0.00	18:30
REFERENCE STANDARD	34.00c	18:30
REFERENCE STANDARD	0.10	18:30
BLANK TEST	0.00	18:31

4/19/22

47A



SUPERVISOR TEST




RAMAPO TOWN
POLICE DEPT
DATAMASTER dmt: 116306
Date: 04/25/2022

OPERATOR NAME:
ANDRES F SANCHEZ
Permit #: 42014
Operator Agency/Dept: RAMAPO TOWN PD
Reference Standard #: 22030

Nominal = 0.10 g/210L

BLANK TEST	0.00	08:36
INTERNAL STANDARD	VERIFIED	08:37
REFERENCE STANDARD	34.00C	
REFERENCE STANDARD	0.10	08:37
BLANK TEST	0.00	08:38
REFERENCE STANDARD	34.00C	
REFERENCE STANDARD	0.10	08:38
BLANK TEST	0.00	08:39
REFERENCE STANDARD	34.00C	
REFERENCE STANDARD	0.10	08:39
BLANK TEST	0.00	08:40

4/25/22
② 47



SUPERVISOR TEST




RAMAPO TOWN
POLICE DEPT
DATAMASTER dmt: 116306
Date: 04/29/2022

OPERATOR NAME:
ANDRES F SANCHEZ
Permit #: 42014
Operator Agency/Dept: RAMAPO TOWN PD
Reference Standard #: 22030

Nominal = 0.10 g/210L

BLANK TEST	0.00	20:25
INTERNAL STANDARD	VERIFIED	20:25
REFERENCE STANDARD	34.01c	
REFERENCE STANDARD	0.10	20:25
BLANK TEST	0.00	20:26
REFERENCE STANDARD	34.00c	
REFERENCE STANDARD	0.10	20:27
BLANK TEST	0.00	20:27
REFERENCE STANDARD	34.00c	
REFERENCE STANDARD	0.10	20:28
REFERENCE STANDARD	0.00	20:28

AS
4/29/22



SOLUTION CHANGE

RAMAPO TOWN
POLICE DEPT
DATAMASTER dmt: 116306




Date: 05/02/2022

OPERATOR NAME:
ANDRES F SANCHEZ
Permit #: 42014
Operator Agency/Dept: RAMAPO TOWN PD
Reference Standard #: 22030

Nominal = 0.10 g/210L

BLANK TEST	0.00	17:44
INTERNAL STANDARD	VERIFIED	17:44
REFERENCE STANDARD	34.00c	
REFERENCE STANDARD	0.10	17:44
BLANK TEST	0.00	17:45
REFERENCE STANDARD	34.01c	
REFERENCE STANDARD	0.10	17:45
BLANK TEST	0.00	17:46
REFERENCE STANDARD	34.01c	
REFERENCE STANDARD	0.10	17:46
BLANK TEST	0.00	17:47

5/2/22
② 47A



SUPERVISOR TEST




RAMAPO TOWN
POLICE DEPT
DATAMASTER dmt: 116306
Date: 05/06/2022

OPERATOR NAME:
ANDRES F SANCHEZ
Permit #: 42014
Operator Agency/Dept: RAMAPO TOWN PD
Reference Standard #: 22030

Nominal = 0.10 g/210L

BLANK TEST	0.00	12:44
INTERNAL STANDARD	VERIFIED	12:44
REFERENCE STANDARD	34.00c	
REFERENCE STANDARD	0.10	12:44
BLANK TEST	0.00	12:45
REFERENCE STANDARD	34.01c	
REFERENCE STANDARD	0.10	12:46
BLANK TEST	0.00	12:46
REFERENCE STANDARD	34.01c	
REFERENCE STANDARD	0.10	12:47
BLANK TEST	0.00	12:48

5/6/22
@ 471



SUPERVISOR TEST




RAMAPO TOWN
POLICE DEPT
DATAMASTER dmt: 116306
Date: 05/10/2022

OPERATOR NAME:
ANDRES F SANCHEZ
Permit #: 42014
Operator Agency/Dept: RAMAPO TOWN PD
Reference Standard #: 22030

Nominal = 0.10 g/210L

BLANK TEST	0.00	16:40
INTERNAL STANDARD	VERIFIED	16:40
REFERENCE STANDARD	34.00c	
REFERENCE STANDARD	0.10	16:41
BLANK TEST	0.00	16:42
REFERENCE STANDARD	34.00c	
REFERENCE STANDARD	0.10	16:42
BLANK TEST	0.00	16:43
REFERENCE STANDARD	34.01c	
REFERENCE STANDARD	0.10	16:43
BLANK TEST	0.00	16:44

5/10/2022
P 474



SUPERVISOR TEST

RAMAPO TOWN
POLICE DEPT
DATAMASTER dmt: 116306
Date: 05/16/2022




OPERATOR NAME:
ANDRES F SANCHEZ
Permit #: 42014
Operator Agency/Dept: RAMAPO TOWN PD
Reference Standard #: 22030

Nominal = 0.10 g/210L

BLANK TEST	0.00	12:04
INTERNAL STANDARD	VERIFIED	12:04
REFERENCE STANDARD	34.00C	
REFERENCE STANDARD	0.10	12:05
BLANK TEST	0.00	12:06
REFERENCE STANDARD	34.00C	
REFERENCE STANDARD	0.10	12:06
BLANK TEST	0.00	12:07
REFERENCE STANDARD	34.01C	
REFERENCE STANDARD	0.10	12:07
REFERENCE STANDARD	0.00	12:08
BLANK TEST		

5/16/2022

2471



SUPERVISOR TEST




RAMAPO TOWN
POLICE DEPT
DATAMASTER dmt: 116306
Date: 05/28/2022

OPERATOR NAME:
ANDRES F SANCHEZ
Permit #: 42014
Operator Agency/Dept: RAMAPO TOWN PD
Reference Standard #: 22030

Nominal = 0.10 g/210L

BLANK TEST	0.00	10:48
INTERNAL STANDARD	VERIFIED	10:48
REFERENCE STANDARD	34.01C	
BLANK TEST	0.10	10:48
REFERENCE STANDARD	0.00	10:49
BLANK TEST	34.00C	
REFERENCE STANDARD	0.10	10:49
REFERENCE STANDARD	0.00	10:50
BLANK TEST	34.00C	
REFERENCE STANDARD	0.10	10:51
REFERENCE STANDARD	0.00	10:51
BLANK TEST		

5/28/2022
471



SOLUTION CHANGE

RAMAPO TOWN
POLICE DEPT
DATAMASTER dmt: 116306
Date: 06/01/2022

OPERATOR NAME:
ANDRES F SANCHEZ
Permit #: 42014
Operator Agency/Dept: RAMAPO TOWN PD
Reference Standard #: 22100


Nominal = 0.10 g/210L

BLANK TEST
INTERNAL STANDARD 0.00 18:52
REFERENCE STANDARD VERIFIED 18:52
REFERENCE STANDARD TEMP 34.00C
REFERENCE STANDARD INTERFERENCE 18:52

Blank Error

6/1/2022
AFS

Alcohol (g/210L)
Flow Rate (L/M)



SOLUTION CHANGE

 RAMAPO TOWN
 POLICE DEPT
 DATAMASTER dmt: 116306



 Date: 06/01/2022

OPERATOR NAME:
 ANDRES F SANCHEZ
 Permit #: 42014
 Operator Agency/Dept: RAMAPO TOWN PD
 Reference Standard #: 22100

Nominal = 0.10 g/210L

BLANK TEST	0.00	18:59
INTERNAL STANDARD	VERIFIED	18:59
REFERENCE STANDARD	34.00C	
REFERENCE STANDARD	0.10	18:59
BLANK TEST	0.00	19:00
REFERENCE STANDARD	34.00C	
REFERENCE STANDARD	0.10	19:00

Blank Error



SUPERVISOR TEST




RAMAPO TOWN
POLICE DEPT
DATAMASTER dmt: 116306
Date: 06/07/2022

OPERATOR NAME:
ANDRES F SANCHEZ
Permit #: 42014
Operator Agency/Dept: RAMAPO TOWN PD
Reference Standard #: 22100

Nominal = 0.10 g/210L

BLANK TEST	0.00	19:13
INTERNAL STANDARD	VERIFIED	19:13
REFERENCE STANDARD	34.00c	
REFERENCE STANDARD	0.10	19:14
BLANK TEST	0.00	19:15
REFERENCE STANDARD	34.00c	
REFERENCE STANDARD	0.10	19:15
BLANK TEST	0.00	19:16
REFERENCE STANDARD	34.00c	
REFERENCE STANDARD	0.10	19:16
BLANK TEST	0.00	19:17

6/7/2022
@V77



SUPERVISOR TEST




RAMAPO TOWN
POLICE DEPT
DATAMASTER dmt: 116306
Date: 06/14/2022

OPERATOR NAME:
ANDRES F SANCHEZ
Permit #: 42014
Operator Agency/Dept: RAMAPO TOWN PD
Reference Standard #: 22100

Nominal = 0.10 g/210L

BLANK TEST	0.00	09:57
INTERNAL STANDARD	VERIFIED	09:57
REFERENCE STANDARD	34.00c	
REFERENCE STANDARD	0.10	09:57
BLANK TEST	0.00	09:58
REFERENCE STANDARD	34.00c	
REFERENCE STANDARD	0.10	09:58
BLANK TEST	0.00	09:59
REFERENCE STANDARD	34.00c	
REFERENCE STANDARD	0.10	10:00
BLANK TEST	0.00	10:00

6/14/22
QMF



SUPERVISOR TEST

RAMAPO TOWN
POLICE DEPT
DATAMASTER dmt: 116306
Date: 06/17/2022




OPERATOR NAME:
ANDRES F SANCHEZ
Permit #: 42014
Operator Agency/Dept: RAMAPO TOWN PD
Reference Standard #: 22100

Nominal = 0.10 g/210L

BLANK TEST	0.00	15:36
INTERNAL STANDARD	VERIFIED	15:36
REFERENCE STANDARD	34.00C	
REFERENCE STANDARD	0.10	15:37
BLANK TEST	0.00	15:38
REFERENCE STANDARD	34.00C	
REFERENCE STANDARD	0.10	15:38
BLANK TEST	0.00	15:39
REFERENCE STANDARD	34.00C	
REFERENCE STANDARD	0.10	15:39
BLANK TEST	0.00	15:40

6/17/22

47



SUPERVISOR TEST




RAMAPO TOWN
POLICE DEPT
DATAMASTER dmt: 116306
Date: 06/27/2022

OPERATOR NAME:
ANDRES F SANCHEZ
Permit #: 42014
Operator Agency/Dept: RAMAPO TOWN PD
Reference Standard #: 22100

Nominal = 0.10 g/210L

BLANK TEST	0.00	10:14
INTERNAL STANDARD	VERIFIED	10:14
REFERENCE STANDARD	34.00C	
REFERENCE STANDARD	0.10	10:15
BLANK TEST	0.00	10:16
REFERENCE STANDARD	34.00C	
REFERENCE STANDARD	0.10	10:16
BLANK TEST	0.00	10:17
REFERENCE STANDARD	34.00C	
REFERENCE STANDARD	0.10	10:17
BLANK TEST	0.00	10:18

6/27/2022
477



SUPERVISOR TEST




RAMAPO TOWN
 POLICE DEPT
 DATAMASTER dmt: 116306
 Date: 06/30/2022

OPERATOR NAME:
 ANDRES F SANCHEZ
 Permit #: 42014
 Operator Agency/Dept: RAMAPO TOWN PD
 Reference Standard #: 22100

Nominal = 0.10 g/210L

BLANK TEST	0.00	15:57
INTERNAL STANDARD	VERIFIED	15:57
REFERENCE STANDARD	34.00C	
REFERENCE STANDARD	0.10	15:58
BLANK TEST	0.00	15:59
REFERENCE STANDARD	34.00C	
REFERENCE STANDARD	0.10	15:59
BLANK TEST	0.00	16:00
REFERENCE STANDARD	34.00C	
REFERENCE STANDARD	0.10	16:00
BLANK TEST	0.00	16:01

6/30/22
 (R) 471



SOLUTION CHANGE

RAMAPO TOWN
POLICE DEPT
DATAMASTER dmt: 116306
Date: 07/06/2022




OPERATOR NAME:
ANDRES F SANCHEZ
Permit #: 42014
Operator Agency/Dept: RAMAPO TOWN PD
Reference Standard #: 22100

Nominal = 0.10 g/210L

BLANK TEST	0.00	14:47
INTERNAL STANDARD	VERIFIED	14:47
REFERENCE STANDARD	34.00c	
REFERENCE STANDARD	0.10	14:47
BLANK TEST	0.00	14:48
REFERENCE STANDARD	34.00c	
REFERENCE STANDARD	0.10	14:49
BLANK TEST	0.00	14:49
REFERENCE STANDARD	34.00c	
REFERENCE STANDARD	0.10	14:50
BLANK TEST	0.00	14:51

7/6/2022


AFH



SUPERVISOR TEST



RAMAPO TOWN
POLICE DEPT
DATAMASTER dmt: 116306

Date: 07/08/2022

OPERATOR NAME:
ANDRES F SANCHEZ
Permit #: 42014
Operator Agency/Dept: RAMAPO TOWN PD
Reference Standard #: 22100

Nominal = 0.10 g/210L

BLANK TEST	0.00	17:39
INTERNAL STANDARD	VERIFIED	17:39
REFERENCE STANDARD	34.00C	
REFERENCE STANDARD	0.10	17:39
REFERENCE STANDARD	0.00	17:40
BLANK TEST	34.00C	
REFERENCE STANDARD	0.10	17:40
REFERENCE STANDARD	0.00	17:41
BLANK TEST	34.00C	
REFERENCE STANDARD	0.10	17:42
REFERENCE STANDARD	0.00	17:42






SUPERVISOR TEST

RAMAPO TOWN
 POLICE DEPT
 DATAMASTER dmt: 116306
 Date: 07/11/2022

OPERATOR NAME:
 JOSEPH T GALVIN
 Permit #: 45745
 Operator Agency/Dept: RAMAPO TOWN PD
 Reference Standard #: 22100

Nominal = 0.10 g/210L

BLANK TEST	0.00	10:29
INTERNAL STANDARD	VERIFIED	10:29
REFERENCE STANDARD	34.00c	
REFERENCE STANDARD	0.10	10:30
BLANK TEST	0.00	10:30
REFERENCE STANDARD	34.01c	
REFERENCE STANDARD	0.10	10:31
BLANK TEST	0.00	10:32
REFERENCE STANDARD	34.00c	
REFERENCE STANDARD	0.10	10:32
BLANK TEST	0.00	10:33



SUPERVISOR TEST




RAMAPO TOWN
POLICE DEPT
DATAMASTER dmt: 116306
Date: 07/18/2022

OPERATOR NAME:
ANDRES F SANCHEZ
Permit #: 42014
Operator Agency/Dept: RAMAPO TOWN PD
Reference Standard #: 22100

Nominal = 0.10 g/210L

BLANK TEST	0.00	08:48
INTERNAL STANDARD	VERIFIED	08:48
REFERENCE STANDARD	34.00c	
REFERENCE STANDARD	0.10	08:49
BLANK TEST	0.00	08:50
REFERENCE STANDARD	34.00c	
REFERENCE STANDARD	0.10	08:50
BLANK TEST	0.00	08:51
REFERENCE STANDARD	34.00c	
REFERENCE STANDARD	0.10	08:51
BLANK TEST	0.00	08:52

7/18/22
@247



SUPERVISOR TEST

RAMAPO TOWN
POLICE DEPT
DATAMASTER dmt: 116306




Date: 07/18/2022

OPERATOR NAME:
ANDRES F SANCHEZ
Permit #: 42014
Operator Agency/Dept: RAMAPO TOWN PD
Reference Standard #: 22100

Nominal = 0.10 g/210L

BLANK TEST	0.00	17:09
INTERNAL STANDARD	VERIFIED	17:09
REFERENCE STANDARD	34.00c	
REFERENCE STANDARD	0.10	17:09
BLANK TEST	0.00	17:10
REFERENCE STANDARD	34.00c	
REFERENCE STANDARD	0.10	17:10
BLANK TEST	0.00	17:11
REFERENCE STANDARD	34.00c	
REFERENCE STANDARD	0.10	17:12
BLANK TEST	0.00	17:12

7/18/22
AS



SUPERVISOR TEST




RAMAPO TOWN
POLICE DEPT
DATAMASTER dmt: 116306
Date: 07/22/2022

OPERATOR NAME:
ANDRES F SANCHEZ
Permit #: 42014
Operator Agency/Dept: RAMAPO TOWN PD
Reference Standard #: 22100

Nominal = 0.10 g/210L

BLANK TEST	0.00	20:59
INTERNAL STANDARD	VERIFIED	20:59
REFERENCE STANDARD	34.00C	
REFERENCE STANDARD	0.10	20:59
BLANK TEST	0.00	21:00
REFERENCE STANDARD	34.00C	
REFERENCE STANDARD	0.10	21:00
BLANK TEST	0.00	21:01
REFERENCE STANDARD	34.00C	
REFERENCE STANDARD	0.10	21:01
BLANK TEST	0.00	21:02

7/22/22
P 47A



SUPERVISOR TEST




RAMAPO TOWN
 POLICE DEPT
 DATAMASTER dmt: 116306
 Date: 07/25/2022

OPERATOR NAME:
 ANDRES F SANCHEZ
 Permit #: 42014
 Operator Agency/Dept: RAMAPO TOWN PD
 Reference Standard #: 22100

Nominal = 0.10 g/210L

BLANK TEST	0.00	09:57
INTERNAL STANDARD	VERIFIED	09:58
REFERENCE STANDARD	34.00C	
REFERENCE STANDARD	0.10	09:58
BLANK TEST	0.00	09:59
REFERENCE STANDARD	34.00C	
REFERENCE STANDARD	0.10	09:59
BLANK TEST	0.00	10:00
REFERENCE STANDARD	34.00C	
REFERENCE STANDARD	0.10	10:00
BLANK TEST	0.00	10:01

7/25/22
Q471



SOLUTION CHANGE

RAMAPO TOWN
POLICE DEPT
DATAMASTER dmt: 116306

Date: 07/29/2022


OPERATOR NAME:
MICHAEL S CARBONE
Permit #: 50609
Operator Agency/Dept: RAMAPO TOWN PD
Reference Standard #: 22190

Nominal = 0.10 g/210L

BLANK TEST 0.00 17:29
INTERNAL STANDARD VERIFIED 17:30
REFERENCE STANDARD TEMP 34.00C
REFERENCE STANDARD 0.09 17:30

Blank Error

7/29/22
MSC



SOLUTION CHANGE




RAMAPO TOWN
POLICE DEPT
DATAMASTER dmt: 116306
Date: 07/29/2022

OPERATOR NAME:
MICHAEL S CARBONE
Permit #: 50609
Operator Agency/Dept: RAMAPO TOWN PD
Reference Standard #: 22190

Nominal = 0.10 g/210L

BLANK TEST	0.00	17:55
INTERNAL STANDARD	VERIFIED	17:55
REFERENCE STANDARD	34.00C	
REFERENCE STANDARD	0.10	17:55
BLANK TEST	0.00	17:56
REFERENCE STANDARD	34.00C	
REFERENCE STANDARD	0.10	17:56
BLANK TEST	0.00	17:57
REFERENCE STANDARD	34.00C	
REFERENCE STANDARD	0.10	17:58
BLANK TEST	0.00	17:58

7/29/22
MC 520



SUPERVISOR TEST




RAMAPO TOWN
POLICE DEPT
DATAMASTER dmt: 116306
Date: 08/02/2022

OPERATOR NAME:
ANDRES F SANCHEZ
Permit #: 42014
Operator Agency/Dept: RAMAPO TOWN PD
Reference Standard #: 22190

Nominal = 0.10 g/210L

BLANK TEST	0.00	09:28
INTERNAL STANDARD	VERIFIED	09:28
REFERENCE STANDARD	34.00C	
REFERENCE STANDARD	0.10	09:29
BLANK TEST	0.00	09:30
REFERENCE STANDARD	34.00C	
REFERENCE STANDARD	0.10	09:30
BLANK TEST	0.00	09:31
REFERENCE STANDARD	34.00C	
REFERENCE STANDARD	0.10	09:31
BLANK TEST	0.00	09:32

8/2/22
AFS



SUPERVISOR TEST




RAMAPO TOWN
POLICE DEPT
DATAMASTER dmt: 116306
Date: 08/11/2022

OPERATOR NAME:
MICHAEL S CARBONE
Permit #: 50609
Operator Agency/Dept: RAMAPO TOWN PD
Reference Standard #: 22190

Nominal = 0.10 g/210L

BLANK TEST	0.00	00:57
INTERNAL STANDARD	VERIFIED	00:57
REFERENCE STANDARD	34.00c	
REFERENCE STANDARD	0.09	00:58
BLANK TEST	0.00	00:59
REFERENCE STANDARD	34.00c	
REFERENCE STANDARD	0.10	00:59
BLANK TEST	0.00	01:00
REFERENCE STANDARD	34.00c	
REFERENCE STANDARD	0.10	01:00
BLANK TEST	0.00	01:01

8/11/22
MC 520



SUPERVISOR TEST




RAMAPO TOWN
POLICE DEPT
DATAMASTER dmt: 116306
Date: 08/17/2022

OPERATOR NAME:
MICHAEL S CARBONE
Permit #: 50609
Operator Agency/Dept: RAMAPO TOWN PD
Reference Standard #: 22190

Nominal = 0.10 g/210L

BLANK TEST	0.00	22:03
INTERNAL STANDARD	VERIFIED	22:03
REFERENCE STANDARD	34.00C	
REFERENCE STANDARD	0.10	22:04
BLANK TEST	0.00	22:05
REFERENCE STANDARD	34.00C	
REFERENCE STANDARD	0.10	22:05
BLANK TEST	0.00	22:06
REFERENCE STANDARD	34.00C	
REFERENCE STANDARD	0.09	22:06
BLANK TEST	0.00	22:07

8/17/22
MC500



SUPERVISOR TEST




RAMAPO TOWN
POLICE DEPT
DATAMASTER dmt: 116306
Date: 08/22/2022

OPERATOR NAME:
MICHAEL S CARBONE
Permit #: 50609
Operator Agency/Dept: RAMAPO TOWN PD
Reference Standard #: 22190

Nominal = 0.10 g/210L

BLANK TEST	0.00	05:16
INTERNAL STANDARD	VERIFIED	05:16
REFERENCE STANDARD	34.00c	
REFERENCE STANDARD	0.09	05:17
BLANK TEST	0.00	05:18
REFERENCE STANDARD	34.00c	
REFERENCE STANDARD	0.09	05:18
REFERENCE STANDARD	0.00	05:19
BLANK TEST	0.00	05:19
REFERENCE STANDARD	34.00c	
REFERENCE STANDARD	0.10	05:19
BLANK TEST	0.00	05:20

8/22/22
MC 570



SUPERVISOR TEST

RAMAPO TOWN
POLICE DEPT
DATAMASTER dmt: 116306

Date: 08/29/2022

OPERATOR NAME:
ANDRES F SANCHEZ
Permit #: 42014
Operator Agency/Dept: RAMAPO TOWN PD
Reference Standard #: 22190

Nominal = 0.10 g/210L

BLANK TEST	0.00	12:33
INTERNAL STANDARD	VERIFIED	12:33
REFERENCE STANDARD	34.00C	
REFERENCE STANDARD	0.09	12:33
BLANK TEST	0.00	12:34
REFERENCE STANDARD	34.00C	
REFERENCE STANDARD	0.09	12:35
BLANK TEST	0.00	12:35
REFERENCE STANDARD	34.00C	
REFERENCE STANDARD	0.09	12:36
BLANK TEST	0.00	12:37

8/29/22
[Signature]

