

SOLUTION CHANGE

RAMAPO TOWN
POLICE DEPT
DATAMASTER dmt: 116306




Date: 01/01/2023

OPERATOR NAME:
MICHAEL CARBONE
Permit #: 50609
Operator Agency/Dept: RAMAPO TOWN PD
Reference Standard #: 22390

Nominal = 0.10 g/210L

| | | |
|-------------------------|----------|-------|
| BLANK TEST | 0.00 | 03:28 |
| INTERNAL STANDARD | VERIFIED | 03:28 |
| REFERENCE STANDARD TEMP | 34.00c | |
| REFERENCE STANDARD | 0.10 | 03:28 |
| BLANK TEST | 0.00 | 03:29 |
| REFERENCE STANDARD TEMP | 34.00c | |
| REFERENCE STANDARD | 0.10 | 03:29 |
| BLANK TEST | 0.00 | 03:30 |
| REFERENCE STANDARD TEMP | 34.00c | |
| REFERENCE STANDARD | 0.10 | 03:31 |
| BLANK TEST | 0.00 | 03:31 |

01/01/23
MC



SOLUTION CHANGE

RAMAPO TOWN
POLICE DEPT
DATAMASTER dmt: 116306

Date: 01/01/2023

OPERATOR NAME:
MICHAEL CARBONE
Permit #: 50609
Operator Agency/Dept: RAMAPO TOWN PD
Reference Standard #: 22390


Nominal = 0.10 g/210L

| | | |
|-------------------------|----------|-------|
| BLANK TEST | 0.00 | 03:02 |
| INTERNAL STANDARD | VERIFIED | 03:02 |
| REFERENCE STANDARD TEMP | 34.00c | |
| REFERENCE STANDARD | 0.09 | 03:02 |

Blank Error

01/01/23

MC 520



SUPERVISOR TEST

RAMAPO TOWN
 POLICE DEPT
 DATAMASTER dmt: 116306

Date: 01/09/2023




OPERATOR NAME:
 MICHAEL S CARBONE
 Permit #: 50609
 Operator Agency/Dept: RAMAPO TOWN PD
 Reference Standard #: 22390

Nominal = 0.10 g/210L

| | | |
|-------------------------|----------|-------|
| BLANK TEST | 0.00 | 05:13 |
| INTERNAL STANDARD | VERIFIED | 05:13 |
| REFERENCE STANDARD TEMP | 34.00c | |
| REFERENCE STANDARD | 0.10 | 05:13 |
| BLANK TEST | 0.00 | 05:14 |
| REFERENCE STANDARD TEMP | 34.00c | |
| REFERENCE STANDARD | 0.10 | 05:14 |
| BLANK TEST | 0.00 | 05:15 |
| REFERENCE STANDARD TEMP | 34.00c | |
| REFERENCE STANDARD | 0.10 | 05:15 |
| BLANK TEST | 0.00 | 05:16 |

01/09/23

MSC



SUPERVISOR TEST

RAMAPO TOWN
POLICE DEPT
DATAMASTER dmt: 116306




Date: 01/19/2023

OPERATOR NAME:
MICHAEL S CARBONE
Permit #: 50609
Operator Agency/Dept: RAMAPO TOWN PD
Reference Standard #: 22390

Nominal = 0.10 g/210L

| | | |
|-------------------------|----------|-------|
| BLANK TEST | 0.00 | 05:57 |
| INTERNAL STANDARD | VERIFIED | 05:57 |
| REFERENCE STANDARD TEMP | 34.00c | |
| REFERENCE STANDARD | 0.10 | 05:58 |
| BLANK TEST | 0.00 | 05:58 |
| REFERENCE STANDARD TEMP | 34.00c | |
| REFERENCE STANDARD | 0.10 | 05:59 |
| BLANK TEST | 0.00 | 06:00 |
| REFERENCE STANDARD TEMP | 34.00c | |
| REFERENCE STANDARD | 0.10 | 06:00 |
| BLANK TEST | 0.00 | 06:01 |

1/19/23



SUPERVISOR TEST

RAMAPO TOWN
POLICE DEPT
DATAMASTER dmt: 116306




Date: 01/26/2023

OPERATOR NAME:
JAMES TESSEYMAN
Permit #: 47783
Operator Agency/Dept: RAMAPO TOWN PD
Reference Standard #: 22390

Nominal = 0.10 g/210L

| | | |
|-------------------------|----------|-------|
| BLANK TEST | 0.00 | 13:13 |
| INTERNAL STANDARD | VERIFIED | 13:13 |
| REFERENCE STANDARD TEMP | 34.00c | |
| REFERENCE STANDARD | 0.09 | 13:13 |
| BLANK TEST | 0.00 | 13:14 |
| REFERENCE STANDARD TEMP | 34.00c | |
| REFERENCE STANDARD | 0.10 | 13:14 |
| BLANK TEST | 0.00 | 13:15 |
| REFERENCE STANDARD TEMP | 34.00c | |
| REFERENCE STANDARD | 0.10 | 13:15 |
| BLANK TEST | 0.00 | 13:16 |

01/26/23
JT492



SUPERVISOR TEST

RAMAPO TOWN
POLICE DEPT
DATAMASTER dmt: 116306




Date: 02/07/2023

OPERATOR NAME:
MICHAEL CARBONE
Permit #: 50609
Operator Agency/Dept: RAMAPO TOWN PD
Reference Standard #: 22390

Nominal = 0.10 g/210L

| | | |
|-------------------------|----------|-------|
| BLANK TEST | 0.00 | 01:07 |
| INTERNAL STANDARD | VERIFIED | 01:07 |
| REFERENCE STANDARD TEMP | 34.00c | |
| REFERENCE STANDARD | 0.10 | 01:07 |
| BLANK TEST | 0.00 | 01:08 |
| REFERENCE STANDARD TEMP | 34.00c | |
| REFERENCE STANDARD | 0.10 | 01:09 |
| BLANK TEST | 0.00 | 01:09 |
| REFERENCE STANDARD TEMP | 34.00c | |
| REFERENCE STANDARD | 0.10 | 01:10 |
| BLANK TEST | 0.00 | 01:11 |

2/7/23
ACF



SOLUTION CHANGE

RAMAPO TOWN
POLICE DEPT
DATAMASTER dmt: 116306




Date: 02/07/2023

OPERATOR NAME:
MICHAEL S CARBONE
Permit #: 50609
Operator Agency/Dept: RAMAPO TOWN PD
Reference Standard #: 22420

Nominal = 0.10 g/210L

| | | |
|-------------------------|----------|-------|
| BLANK TEST | 0.00 | 11:43 |
| INTERNAL STANDARD | VERIFIED | 11:43 |
| REFERENCE STANDARD TEMP | 34.00c | |
| REFERENCE STANDARD | 0.10 | 11:44 |
| BLANK TEST | 0.00 | 11:45 |
| REFERENCE STANDARD TEMP | 34.00c | |
| REFERENCE STANDARD | 0.10 | 11:45 |
| BLANK TEST | 0.00 | 11:46 |
| REFERENCE STANDARD TEMP | 34.00c | |
| REFERENCE STANDARD | 0.10 | 11:46 |
| BLANK TEST | 0.00 | 11:47 |

2/7/23
MC 520



SUPERVISOR TEST

RAMAPO TOWN
POLICE DEPT
DATAMASTER dmt: 116306




Date: 02/17/2023

OPERATOR NAME:
MICHAEL S CARBONE
Permit #: 50609
Operator Agency/Dept: RAMAPO TOWN PD
Reference Standard #: 22420

Nominal = 0.10 g/210L

| | | |
|-------------------------|----------|-------|
| BLANK TEST | 0.00 | 00:32 |
| INTERNAL STANDARD | VERIFIED | 00:32 |
| REFERENCE STANDARD TEMP | 34.00c | |
| REFERENCE STANDARD | 0.10 | 00:32 |
| BLANK TEST | 0.00 | 00:33 |
| REFERENCE STANDARD TEMP | 34.00c | |
| REFERENCE STANDARD | 0.10 | 00:33 |
| BLANK TEST | 0.00 | 00:34 |
| REFERENCE STANDARD TEMP | 34.00c | |
| REFERENCE STANDARD | 0.10 | 00:34 |
| BLANK TEST | 0.00 | 00:35 |

2/17/23
AC520



SOLUTION CHANGE

RAMAPO TOWN
POLICE DEPT
DATAMASTER dmt: 116306




Date: 03/01/2023

OPERATOR NAME:
MICHAEL S CARBONE
Permit #: 50609
Operator Agency/Dept: RAMAPO TOWN PD
Reference Standard #: 22420

Nominal = 0.10 g/210L

| | | |
|-------------------------|----------|-------|
| BLANK TEST | 0.00 | 04:26 |
| INTERNAL STANDARD | VERIFIED | 04:27 |
| REFERENCE STANDARD TEMP | 34.00c | |
| REFERENCE STANDARD | 0.10 | 04:27 |
| BLANK TEST | 0.00 | 04:28 |
| REFERENCE STANDARD TEMP | 34.00c | |
| REFERENCE STANDARD | 0.10 | 04:28 |
| BLANK TEST | 0.00 | 04:29 |
| REFERENCE STANDARD TEMP | 34.00c | |
| REFERENCE STANDARD | 0.10 | 04:29 |
| BLANK TEST | 0.00 | 04:30 |

3/1/23
MSC



SUPERVISOR TEST

RAMAPO TOWN
POLICE DEPT
DATAMASTER dmt: 116306




Date: 03/09/2023

OPERATOR NAME:
MICHAEL S CARBONE
Permit #: 50609
Operator Agency/Dept: RAMAPO TOWN PD
Reference Standard #: 22420

Nominal = 0.10 g/210L

| | | |
|-------------------------|----------|-------|
| BLANK TEST | 0.00 | 01:05 |
| INTERNAL STANDARD | VERIFIED | 01:05 |
| REFERENCE STANDARD TEMP | 34.00c | |
| REFERENCE STANDARD | 0.10 | 01:06 |
| BLANK TEST | 0.00 | 01:07 |
| REFERENCE STANDARD TEMP | 34.00c | |
| REFERENCE STANDARD | 0.10 | 01:07 |
| BLANK TEST | 0.00 | 01:08 |
| REFERENCE STANDARD TEMP | 34.00c | |
| REFERENCE STANDARD | 0.10 | 01:08 |
| BLANK TEST | 0.00 | 01:09 |

3/9/23
Me800



SUPERVISOR TEST

RAMAPO TOWN
 POLICE DEPT
 DATAMASTER dmt: 116306




Date: 03/19/2023

OPERATOR NAME:
 MICHAEL S CARBONE
 Permit #: 50609
 Operator Agency/Dept: RAMAPO TOWN PD
 Reference Standard #: 22420

Nominal = 0.10 g/210L

| | | |
|-------------------------|----------|-------|
| BLANK TEST | 0.00 | 01:21 |
| INTERNAL STANDARD | VERIFIED | 01:21 |
| REFERENCE STANDARD TEMP | 34.00c | |
| REFERENCE STANDARD | 0.10 | 01:21 |
| BLANK TEST | 0.00 | 01:22 |
| REFERENCE STANDARD TEMP | 34.00c | |
| REFERENCE STANDARD | 0.10 | 01:22 |
| BLANK TEST | 0.00 | 01:23 |
| REFERENCE STANDARD TEMP | 34.00c | |
| REFERENCE STANDARD | 0.10 | 01:23 |
| BLANK TEST | 0.00 | 01:24 |

3/19/23
MC520



SOLUTION CHANGE

RAMAPO TOWN
POLICE DEPT
DATAMASTER dmt: 116306




Date: 04/01/2023

OPERATOR NAME:
MICHAEL S CARBONE
Permit #: 50609
Operator Agency/Dept: RAMAPO TOWN PD
Reference Standard #: 23030

Nominal = 0.10 g/210L

| | | |
|-------------------------|----------|-------|
| BLANK TEST | 0.00 | 01:15 |
| INTERNAL STANDARD | VERIFIED | 01:15 |
| REFERENCE STANDARD TEMP | 34.00c | |
| REFERENCE STANDARD | 0.10 | 01:16 |
| BLANK TEST | 0.00 | 01:16 |
| REFERENCE STANDARD TEMP | 34.00c | |
| REFERENCE STANDARD | 0.10 | 01:17 |
| BLANK TEST | 0.00 | 01:18 |
| REFERENCE STANDARD TEMP | 34.00c | |
| REFERENCE STANDARD | 0.10 | 01:18 |
| BLANK TEST | 0.00 | 01:19 |

4/1/23
MC 520



SUPERVISOR TEST

RAMAPO TOWN
POLICE DEPT
DATAMASTER dmt: 116306

Date: 04/08/2023




OPERATOR NAME:
MICHAEL S CARBONE
Permit #: 50609
Operator Agency/Dept: RAMAPO TOWN PD
Reference Standard #: 23030

Nominal = 0.10 g/210L

| | | |
|-------------------------|----------|-------|
| BLANK TEST | 0.00 | 23:28 |
| INTERNAL STANDARD | VERIFIED | 23:28 |
| REFERENCE STANDARD TEMP | 34.00c | |
| REFERENCE STANDARD | 0.10 | 23:28 |
| BLANK TEST | 0.00 | 23:29 |
| REFERENCE STANDARD TEMP | 34.00c | |
| REFERENCE STANDARD | 0.10 | 23:30 |
| BLANK TEST | 0.00 | 23:30 |
| REFERENCE STANDARD TEMP | 34.00c | |
| REFERENCE STANDARD | 0.10 | 23:31 |
| BLANK TEST | 0.00 | 23:32 |

4/8/2023

MC520



SUPERVISOR TEST

RAMAPO TOWN
POLICE DEPT
DATAMASTER dmt: 116306




Date: 04/13/2023

OPERATOR NAME:
ANDRES F SANCHEZ
Permit #: 42014
Operator Agency/Dept: RAMAPO TOWN PD
Reference Standard #: 23030

Nominal = 0.10 g/210L

| | | |
|-------------------------|----------|-------|
| BLANK TEST | 0.00 | 15:36 |
| INTERNAL STANDARD | VERIFIED | 15:36 |
| REFERENCE STANDARD TEMP | 34.00c | |
| REFERENCE STANDARD | 0.09 | 15:36 |
| BLANK TEST | 0.00 | 15:37 |
| REFERENCE STANDARD TEMP | 34.00c | |
| REFERENCE STANDARD | 0.10 | 15:37 |
| BLANK TEST | 0.00 | 15:38 |
| REFERENCE STANDARD TEMP | 34.00c | |
| REFERENCE STANDARD | 0.10 | 15:39 |
| BLANK TEST | 0.00 | 15:39 |

4/13/2023
②471



SUPERVISOR TEST

RAMAPO TOWN
POLICE DEPT
DATAMASTER dmt: 116306

Date: 04/23/2023



OPERATOR NAME:
MICHAEL S CARBONE
Permit #: 50609
Operator Agency/Dept: RAMAPO TOWN PD
Reference Standard #: 23030

Nominal = 0.10 g/210L

| | | |
|-------------------------|----------|-------|
| BLANK TEST | 0.00 | 05:51 |
| INTERNAL STANDARD | VERIFIED | 05:51 |
| REFERENCE STANDARD TEMP | 34.00c | |
| REFERENCE STANDARD | 0.10 | 05:51 |
| BLANK TEST | 0.00 | 05:52 |
| REFERENCE STANDARD TEMP | 34.00c | |
| REFERENCE STANDARD | 0.10 | 05:52 |
| BLANK TEST | 0.00 | 05:53 |
| REFERENCE STANDARD TEMP | 34.00c | |
| REFERENCE STANDARD | 0.10 | 05:53 |
| BLANK TEST | 0.00 | 05:54 |

4/23/23

Mc520



SUPERVISOR TEST

RAMAPO TOWN
 POLICE DEPT
 DATAMASTER dmt: 116306




Date: 04/29/2023

OPERATOR NAME:
 MICHAEL S CARBONE
 Permit #: 50609
 Operator Agency/Dept: RAMAPO TOWN PD
 Reference Standard #: 23030

Nominal = 0.10 g/210L

| | | |
|-------------------------|----------|-------|
| BLANK TEST | 0.00 | 23:37 |
| INTERNAL STANDARD | VERIFIED | 23:37 |
| REFERENCE STANDARD TEMP | 34.00c | |
| REFERENCE STANDARD | 0.10 | 23:37 |
| BLANK TEST | 0.00 | 23:38 |
| REFERENCE STANDARD TEMP | 34.01c | |
| REFERENCE STANDARD | 0.10 | 23:38 |
| BLANK TEST | 0.00 | 23:39 |
| REFERENCE STANDARD TEMP | 34.01c | |
| REFERENCE STANDARD | 0.10 | 23:39 |
| BLANK TEST | 0.00 | 23:40 |

4/29/23
M/C



SOLUTION CHANGE

RAMAPO TOWN
POLICE DEPT
DATAMASTER dmt: 116306




Date: 05/02/2023

OPERATOR NAME:
ANDRES F SANCHEZ
Permit #: 42014
Operator Agency/Dept: RAMAPO TOWN PD
Reference Standard #: 23030

Nominal = 0.10 g/210L

| | | |
|-------------------------|----------|-------|
| BLANK TEST | 0.00 | 17:03 |
| INTERNAL STANDARD | VERIFIED | 17:03 |
| REFERENCE STANDARD TEMP | 34.01c | |
| REFERENCE STANDARD | 0.10 | 17:04 |
| BLANK TEST | 0.00 | 17:05 |
| REFERENCE STANDARD TEMP | 34.00c | |
| REFERENCE STANDARD | 0.10 | 17:05 |
| BLANK TEST | 0.00 | 17:06 |
| REFERENCE STANDARD TEMP | 34.00c | |
| REFERENCE STANDARD | 0.10 | 17:06 |
| BLANK TEST | 0.00 | 17:07 |

5/2/23
①471



SUPERVISOR TEST

RAMAPO TOWN
POLICE DEPT
DATAMASTER dmt: 116306




Date: 05/09/2023

OPERATOR NAME:
MICHAEL S CARBONE
Permit #: 50609
Operator Agency/Dept: RAMAPO TOWN PD
Reference Standard #: 23030

Nominal = 0.10 g/210L

| | | |
|-------------------------|----------|-------|
| BLANK TEST | 0.00 | 00:46 |
| INTERNAL STANDARD | VERIFIED | 00:46 |
| REFERENCE STANDARD TEMP | 33.99c | |
| REFERENCE STANDARD | 0.09 | 00:47 |
| BLANK TEST | 0.00 | 00:47 |
| REFERENCE STANDARD TEMP | 34.00c | |
| REFERENCE STANDARD | 0.10 | 00:48 |
| BLANK TEST | 0.00 | 00:49 |
| REFERENCE STANDARD TEMP | 34.00c | |
| REFERENCE STANDARD | 0.10 | 00:49 |
| BLANK TEST | 0.00 | 00:50 |

Mc520
5/9/23



SUPERVISOR TEST

RAMAPO TOWN
POLICE DEPT
DATAMASTER dmt: 116306




Date: 05/18/2023

OPERATOR NAME:
MICHAEL S CARBONE
Permit #: 50609
Operator Agency/Dept: RAMAPO TOWN PD
Reference Standard #: 23030

Nominal = 0.10 g/210L

| | | |
|-------------------------|----------|-------|
| BLANK TEST | 0.00 | 20:43 |
| INTERNAL STANDARD | VERIFIED | 20:43 |
| REFERENCE STANDARD TEMP | 34.00c | |
| REFERENCE STANDARD | 0.10 | 20:44 |
| BLANK TEST | 0.00 | 20:45 |
| REFERENCE STANDARD TEMP | 34.00c | |
| REFERENCE STANDARD | 0.10 | 20:45 |
| BLANK TEST | 0.00 | 20:46 |
| REFERENCE STANDARD TEMP | 34.01c | |
| REFERENCE STANDARD | 0.10 | 20:46 |
| BLANK TEST | 0.00 | 20:47 |

5/18/23
MC520



SUPERVISOR TEST

RAMAPO TOWN
POLICE DEPT
DATAMASTER dmt: 116306




Date: 05/22/2023

OPERATOR NAME:
MICHAEL S CARBONE
Permit #: 50609
Operator Agency/Dept: RAMAPO TOWN PD
Reference Standard #: 23030

Nominal = 0.10 g/210L

| | | |
|-------------------------|----------|-------|
| BLANK TEST | 0.00 | 23:37 |
| INTERNAL STANDARD | VERIFIED | 23:37 |
| REFERENCE STANDARD TEMP | 34.01c | |
| REFERENCE STANDARD | 0.09 | 23:38 |
| BLANK TEST | 0.00 | 23:39 |
| REFERENCE STANDARD TEMP | 34.01c | |
| REFERENCE STANDARD | 0.10 | 23:39 |
| BLANK TEST | 0.00 | 23:40 |
| REFERENCE STANDARD TEMP | 34.00c | |
| REFERENCE STANDARD | 0.10 | 23:40 |
| BLANK TEST | 0.00 | 23:41 |

5/22/23
MCB



SUPERVISOR TEST

RAMAPO TOWN
POLICE DEPT
DATAMASTER dmt: 116306

Date: 05/30/2023




OPERATOR NAME:
MICHAEL S CARBONE
Permit #: 50609
Operator Agency/Dept: RAMAPO TOWN PD
Reference Standard #: 23030

Nominal = 0.10 g/210L

| | | |
|-------------------------|----------|-------|
| BLANK TEST | 0.00 | 01:12 |
| INTERNAL STANDARD | VERIFIED | 01:12 |
| REFERENCE STANDARD TEMP | 34.00c | |
| REFERENCE STANDARD | 0.10 | 01:13 |
| BLANK TEST | 0.00 | 01:14 |
| REFERENCE STANDARD TEMP | 34.00c | |
| REFERENCE STANDARD | 0.10 | 01:14 |
| BLANK TEST | 0.00 | 01:15 |
| REFERENCE STANDARD TEMP | 34.01c | |
| REFERENCE STANDARD | 0.10 | 01:15 |
| BLANK TEST | 0.00 | 01:16 |

05/30/23

MC520



SOLUTION CHANGE

RAMAPO TOWN
POLICE DEPT
DATAMASTER dmt: 116306




Date: 06/07/2023

OPERATOR NAME:
MICHAEL S CARBONE
Permit #: 50609
Operator Agency/Dept: RAMAPO TOWN PD
Reference Standard #: 23030

Nominal = 0.10 g/210L

| | | |
|-------------------------|----------|-------|
| BLANK TEST | 0.00 | 04:00 |
| INTERNAL STANDARD | VERIFIED | 04:01 |
| REFERENCE STANDARD TEMP | 34.00c | |
| REFERENCE STANDARD | 0.10 | 04:01 |
| BLANK TEST | 0.00 | 04:02 |
| REFERENCE STANDARD TEMP | 34.00c | |
| REFERENCE STANDARD | 0.10 | 04:02 |
| BLANK TEST | 0.00 | 04:03 |
| REFERENCE STANDARD TEMP | 34.00c | |
| REFERENCE STANDARD | 0.10 | 04:03 |
| BLANK TEST | 0.00 | 04:04 |

6/7/23
M. Carbone



SUPERVISOR TEST

RAMAPO TOWN
POLICE DEPT
DATAMASTER dmt: 116306




Date: 06/13/2023

OPERATOR NAME:
JAMES TESSEYMAN
Permit #: 47783
Operator Agency/Dept: RAMAPO TOWN PD
Reference Standard #: 23030

Nominal = 0.10 g/210L

| | | |
|-------------------------|----------|-------|
| BLANK TEST | 0.00 | 18:07 |
| INTERNAL STANDARD | VERIFIED | 18:08 |
| REFERENCE STANDARD TEMP | 34.00c | |
| REFERENCE STANDARD | 0.10 | 18:08 |
| BLANK TEST | 0.00 | 18:09 |
| REFERENCE STANDARD TEMP | 34.00c | |
| REFERENCE STANDARD | 0.10 | 18:09 |
| BLANK TEST | 0.00 | 18:10 |
| REFERENCE STANDARD TEMP | 34.00c | |
| REFERENCE STANDARD | 0.10 | 18:10 |
| BLANK TEST | 0.00 | 18:11 |

06/13/2023
JTS



SUPERVISOR TEST

RAMAPO TOWN
POLICE DEPT
DATAMASTER dmt: 116306




Date: 06/23/2023

OPERATOR NAME:
MICHAEL S CARBONE
Permit #: 50609
Operator Agency/Dept: RAMAPO TOWN PD
Reference Standard #: 23030

Nominal = 0.10 g/210L

| | | |
|-------------------------|----------|-------|
| BLANK TEST | 0.00 | 01:14 |
| INTERNAL STANDARD | VERIFIED | 01:14 |
| REFERENCE STANDARD TEMP | 34.00c | |
| REFERENCE STANDARD | 0.10 | 01:15 |
| BLANK TEST | 0.00 | 01:15 |
| REFERENCE STANDARD TEMP | 34.00c | |
| REFERENCE STANDARD | 0.10 | 01:16 |
| BLANK TEST | 0.00 | 01:17 |
| REFERENCE STANDARD TEMP | 34.00c | |
| REFERENCE STANDARD | 0.10 | 01:17 |
| BLANK TEST | 0.00 | 01:18 |

McCar
6/23/23



SUPERVISOR TEST

RAMAPO TOWN
POLICE DEPT
DATAMASTER dmt: 116306

Date: 06/29/2023

OPERATOR NAME:
MICHAEL S CARBONE
Permit #: 50609
Operator Agency/Dept: RAMAPO TOWN PD
Reference Standard #: 23030

Nominal = 0.10 g/210L

| | | |
|-------------------------|----------|-------|
| BLANK TEST | 0.00 | 01:11 |
| INTERNAL STANDARD | VERIFIED | 01:11 |
| REFERENCE STANDARD TEMP | 34.00c | |
| REFERENCE STANDARD | 0.10 | 01:11 |
| BLANK TEST | 0.00 | 01:12 |
| REFERENCE STANDARD TEMP | 34.00c | |
| REFERENCE STANDARD | 0.10 | 01:13 |
| BLANK TEST | 0.00 | 01:13 |
| REFERENCE STANDARD TEMP | 34.00c | |
| REFERENCE STANDARD | 0.10 | 01:14 |
| BLANK TEST | 0.00 | 01:14 |

6/29/23
MCS

