

SOLUTION CHANGE

RAMAPO TOWN
POLICE DEPT
DATAMASTER dmt: 116306

Date: 07/01/2023




OPERATOR NAME:
MICHAEL S CARBONE
Permit #: 50609
Operator Agency/Dept: RAMAPO TOWN PD
Reference Standard #: 23090

Nominal = 0.10 g/210L

BLANK TEST	0.00	05:50
INTERNAL STANDARD	VERIFIED	05:50
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	05:51
BLANK TEST	0.00	05:51
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	05:52
BLANK TEST	0.00	05:53
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	05:53
BLANK TEST	0.00	05:54

7/1/23

McS20



SUPERVISOR TEST

RAMAPO TOWN
 POLICE DEPT
 DATAMASTER dmt: 116306




Date: 07/09/2023

OPERATOR NAME:
 MICHAEL S CARBONE
 Permit #: 50609
 Operator Agency/Dept: RAMAPO TOWN PD
 Reference Standard #: 23090

Nominal = 0.10 g/210L

BLANK TEST	0.00	02:42
INTERNAL STANDARD	VERIFIED	02:43
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	02:43
BLANK TEST	0.00	02:44
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	02:44
BLANK TEST	0.00	02:45
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	02:45
BLANK TEST	0.00	02:46

7/9/23
McE



SUPERVISOR TEST

RAMAPO TOWN
 POLICE DEPT
 DATAMASTER dmt: 116306




Date: 07/18/2023

OPERATOR NAME:
 MICHAEL S CARBONE
 Permit #: 50609
 Operator Agency/Dept: RAMAPO TOWN PD
 Reference Standard #: 23090

Nominal = 0.10 g/210L

BLANK TEST	0.00	13:17
INTERNAL STANDARD	VERIFIED	13:17
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	13:17
BLANK TEST	0.00	13:18
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	13:18
BLANK TEST	0.00	13:19
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	13:19
BLANK TEST	0.00	13:20

McF20
7/18/23



SUPERVISOR TEST

RAMAPO TOWN
POLICE DEPT
DATAMASTER dmt: 116306

Date: 07/29/2023




OPERATOR NAME:
MICHAEL S CARBONE
Permit #: 50609
Operator Agency/Dept: RAMAPO TOWN PD
Reference Standard #: 23090

Nominal = 0.10 g/210L

BLANK TEST	0.00	23:16
INTERNAL STANDARD	VERIFIED	23:17
REFERENCE STANDARD	TEMP	34.00c
REFERENCE STANDARD	0.10	23:17
BLANK TEST	0.00	23:18
REFERENCE STANDARD	TEMP	34.00c
REFERENCE STANDARD	0.10	23:18
BLANK TEST	0.00	23:19
REFERENCE STANDARD	TEMP	34.00c
REFERENCE STANDARD	0.10	23:19
BLANK TEST	0.00	23:20

7/29/23

McCarone



SOLUTION CHANGE

RAMAPO TOWN
POLICE DEPT
DATAMASTER dmt: 116306




Date: 08/06/2023

OPERATOR NAME:
MICHAEL S CARBONE
Permit #: 50609
Operator Agency/Dept: RAMAPO TOWN PD
Reference Standard #: 23090

Nominal = 0.10 g/210L

BLANK TEST	0.00	03:04
INTERNAL STANDARD	VERIFIED	03:04
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	03:05
BLANK TEST	0.00	03:05
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	03:06
BLANK TEST	0.00	03:07
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	03:07
BLANK TEST	0.00	03:08

8/6/23
McCar



SUPERVISOR TEST

RAMAPO TOWN
POLICE DEPT
DATAMASTER dmt: 116306




Date: 08/13/2023

OPERATOR NAME:
MICHAEL S CARBONE
Permit #: 50609
Operator Agency/Dept: RAMAPO TOWN PD
Reference Standard #: 23090

Nominal = 0.10 g/210L

BLANK TEST	0.00	02:53
INTERNAL STANDARD	VERIFIED	02:53
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	02:54
BLANK TEST	0.00	02:54
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	02:55
BLANK TEST	0.00	02:56
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	02:56
BLANK TEST	0.00	02:57

8/13/23
Mc20



SUPERVISOR TEST

RAMAPO TOWN
 POLICE DEPT
 DATAMASTER dmt: 116306




Date: 08/21/2023

OPERATOR NAME:
 JAMES TESSEYMAN
 Permit #: 47783
 Operator Agency/Dept: RAMAPO TOWN PD
 Reference Standard #: 23090

Nominal = 0.10 g/210L

BLANK TEST	0.00	16:16
INTERNAL STANDARD	VERIFIED	16:16
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	16:16
BLANK TEST	0.00	16:17
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	16:18
BLANK TEST	0.00	16:18
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	16:19
BLANK TEST	0.00	16:20

08/21/2023
 JT 492



SUPERVISOR TEST

RAMAPO TOWN
POLICE DEPT
DATAMASTER dmt: 116306




Date: 09/01/2023

OPERATOR NAME:
MICHAEL S CARBONE
Permit #: 50609
Operator Agency/Dept: RAMAPO TOWN PD
Reference Standard #: 23090

Nominal = 0.10 g/210L

BLANK TEST	0.00	01:35
INTERNAL STANDARD	VERIFIED	01:35
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	01:36
BLANK TEST	0.00	01:36
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	01:37
BLANK TEST	0.00	01:37
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	01:38
BLANK TEST	0.00	01:39

9/1/23
MCS



SOLUTION CHANGE

RAMAPO TOWN
 POLICE DEPT
 DATAMASTER dmt: 116306




Date: 09/02/2023

OPERATOR NAME:
 MICHAEL S CARBONE
 Permit #: 50609
 Operator Agency/Dept: RAMAPO TOWN PD
 Reference Standard #: 23190

Nominal = 0.10 g/210L

BLANK TEST	0.00	06:27
INTERNAL STANDARD	VERIFIED	06:27
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	06:27
BLANK TEST	0.00	06:28
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	06:29
BLANK TEST	0.00	06:30
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	06:30
BLANK TEST	0.00	06:31

9/2/23
MC520



SOLUTION CHANGE

RAMAPO TOWN
POLICE DEPT
DATAMASTER dmt: 116306




Date: 09/07/2023

OPERATOR NAME:
ANDRES F SANCHEZ
Permit #: 42014
Operator Agency/Dept: RAMAPO TOWN PD
Reference Standard #: 23190

Nominal = 0.10 g/210L

BLANK TEST	0.00	20:51
INTERNAL STANDARD	VERIFIED	20:51
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	20:51
BLANK TEST	0.00	20:52
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	20:53
BLANK TEST	0.00	20:53
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	20:54
BLANK TEST	0.00	20:55

9/7/23
471



SUPERVISOR TEST

RAMAPO TOWN
POLICE DEPT
DATAMASTER dmt: 116306




Date: 09/20/2023

OPERATOR NAME:
MICHAEL S CARBONE
Permit #: 50609
Operator Agency/Dept: RAMAPO TOWN PD
Reference Standard #: 23190

Nominal = 0.10 g/210L

BLANK TEST	0.00	00:12
INTERNAL STANDARD	VERIFIED	00:12
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	00:12
BLANK TEST	0.00	00:13
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	00:14
BLANK TEST	0.00	00:14
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	00:15
BLANK TEST	0.00	00:15

9/20/23
MCS



SUPERVISOR TEST

RAMAPO TOWN
POLICE DEPT
DATAMASTER dmt: 116306




Date: 09/29/2023

OPERATOR NAME:
MICHAEL S CARBONE
Permit #: 50609
Operator Agency/Dept: RAMAPO TOWN PD
Reference Standard #: 23190

Nominal = 0.10 g/210L

BLANK TEST	0.00	05:28
INTERNAL STANDARD	VERIFIED	05:28
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	05:29
BLANK TEST	0.00	05:29
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	05:30
BLANK TEST	0.00	05:31
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	05:31
BLANK TEST	0.00	05:32

9/29/23
MCS



SOLUTION CHANGE

RAMAPO TOWN
POLICE DEPT
DATAMASTER dmt: 116306

Date: 10/05/2023


OPERATOR NAME:
MICHAEL S CARBONE
Permit #: 50609
Operator Agency/Dept: RAMAPO TOWN PD
Reference Standard #: 23190

Nominal = 0.10 g/210L

BLANK TEST	0.00	04:03
INTERNAL STANDARD	VERIFIED	04:03
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	04:03
BLANK TEST	0.00	04:04
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	04:05
BLANK TEST	0.00	04:05
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	04:06
BLANK TEST	0.00	04:07

10/5/23

MCS



SUPERVISOR TEST

RAMAPO TOWN
 POLICE DEPT
 DATAMASTER dmt: 116306

Date: 10/14/2023




OPERATOR NAME:
 MICHAEL S CARBONE
 Permit #: 50609
 Operator Agency/Dept: RAMAPO TOWN PD
 Reference Standard #: 23190

Nominal = 0.10 g/210L

BLANK TEST	0.00	15:46
INTERNAL STANDARD	VERIFIED	15:47
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	15:47
BLANK TEST	0.00	15:48
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	15:48
BLANK TEST	0.00	15:49
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	15:49
BLANK TEST	0.00	15:50

10/14/23

McS



SUPERVISOR TEST

RAMAPO TOWN
POLICE DEPT
DATAMASTER dmt: 116306

Date: 10/24/2023




OPERATOR NAME:
JAMES TESSEYMAN
Permit #: 47783
Operator Agency/Dept: RAMAPO TOWN PD
Reference Standard #: 23190

Nominal = 0.10 g/210L

BLANK TEST	0.00	11:45
INTERNAL STANDARD	VERIFIED	11:45
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	11:46
BLANK TEST	0.00	11:46
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	11:47
BLANK TEST	0.00	11:48
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	11:48
BLANK TEST	0.00	11:49

10/24/2023

[Handwritten signature]



SOLUTION CHANGE

RAMAPO TOWN
POLICE DEPT
DATAMASTER dmt: 116306




Date: 11/05/2023

OPERATOR NAME:
MICHAEL S CARBONE
Permit #: 50609
Operator Agency/Dept: RAMAPO TOWN PD
Reference Standard #: 23270

Nominal = 0.10 g/210L

BLANK TEST	0.00	06:36
INTERNAL STANDARD	VERIFIED	06:36
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	06:37
BLANK TEST	0.00	06:38
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	06:38
BLANK TEST	0.00	06:39
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	06:39
BLANK TEST	0.00	06:40

11/5/23



SUPERVISOR TEST

RAMAPO TOWN
POLICE DEPT
DATAMASTER dmt: 116306

Date: 11/14/2023




OPERATOR NAME:
MICHAEL S CARBONE
Permit #: 50609
Operator Agency/Dept: RAMAPO TOWN PD
Reference Standard #: 23270

Nominal = 0.10 g/210L

BLANK TEST	0.00	01:11
INTERNAL STANDARD	VERIFIED	01:11
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	01:12
BLANK TEST	0.00	01:12
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	01:13
BLANK TEST	0.00	01:13
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	01:14
BLANK TEST	0.00	01:15

11/14/23

Mc520



SUPERVISOR TEST

RAMAPO TOWN
POLICE DEPT
DATAMASTER dmt: 116306




Date: 11/21/2023

OPERATOR NAME:
MICHAEL S CARBONE
Permit #: 50609
Operator Agency/Dept: RAMAPO TOWN PD
Reference Standard #: 23270

Nominal = 0.10 g/210L

BLANK TEST	0.00	23:34
INTERNAL STANDARD	VERIFIED	23:34
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	23:34
BLANK TEST	0.00	23:35
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	23:35
BLANK TEST	0.00	23:36
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	23:36
BLANK TEST	0.00	23:37

11/21/23
MS



SOLUTION CHANGE

RAMAPO TOWN
POLICE DEPT
DATAMASTER dmt: 116306




Date: 12/05/2023

OPERATOR NAME:
MICHAEL S CARBONE
Permit #: 50609
Operator Agency/Dept: RAMAPO TOWN PD
Reference Standard #: 23270

Nominal = 0.10 g/210L

BLANK TEST	0.00	00:36
INTERNAL STANDARD	VERIFIED	00:36
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	00:37
BLANK TEST	0.00	00:38
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	00:38
BLANK TEST	0.00	00:39
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	00:39
BLANK TEST	0.00	00:40

12/5/23
MSC



SUPERVISOR TEST

RAMAPO TOWN
POLICE DEPT
DATAMASTER dmt: 116306




Date: 12/13/2023

OPERATOR NAME:
MICHAEL S CARBONE
Permit #: 50609
Operator Agency/Dept: RAMAPO TOWN PD
Reference Standard #: 23270

Nominal = 0.10 g/210L

BLANK TEST	0.00	00:42
INTERNAL STANDARD	VERIFIED	00:42
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	00:42
BLANK TEST	0.00	00:43
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	00:43
BLANK TEST	0.00	00:44
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	00:44
BLANK TEST	0.00	00:45

12/13/23
MC



SUPERVISOR TEST

RAMAPO TOWN
POLICE DEPT
DATAMASTER dmt: 116306

Date: 12/20/2023




OPERATOR NAME:
MICHAEL S CARBONE
Permit #: 50609
Operator Agency/Dept: RAMAPO TOWN PD
Reference Standard #: 23270

Nominal = 0.10 g/210L

BLANK TEST	0.00	05:44
INTERNAL STANDARD	VERIFIED	05:45
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	05:45
BLANK TEST	0.00	05:46
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	05:46
BLANK TEST	0.00	05:47
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	05:47
BLANK TEST	0.00	05:48

12/20/23

MSC



SUPERVISOR TEST

RAMAPO TOWN
POLICE DEPT
DATAMASTER dmt: 116306

Date: 12/30/2023

OPERATOR NAME:
MICHAEL S CARBONE
Permit #: 50609
Operator Agency/Dept: RAMAPO TOWN PD
Reference Standard #: 23270

Nominal = 0.10 g/210L

BLANK TEST	0.00	05:52
INTERNAL STANDARD	VERIFIED	05:52
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	05:52
BLANK TEST	0.00	05:53
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	05:53
BLANK TEST	0.00	05:54
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	05:54
BLANK TEST	0.00	05:55

12/30/23

MC520

