

SOLUTION CHANGE

RAMAPO TOWN
POLICE DEPT
DATAMASTER dmt: 116306

Date: 01/03/2024




OPERATOR NAME:
MICHAEL S CARBONE
Permit #: 50609
Operator Agency/Dept: RAMAPO TOWN PD
Reference Standard #: 23350

Nominal = 0.10 g/210L

BLANK TEST	0.00	00:50
INTERNAL STANDARD	VERIFIED	00:51
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	00:51
BLANK TEST	0.00	00:52
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	00:52
BLANK TEST	0.00	00:53
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	00:53
BLANK TEST	0.00	00:54

1/3/24

MS 20



SUPERVISOR TEST

RAMAPO TOWN
POLICE DEPT
DATAMASTER dmt: 116306




Date: 01/11/2024

OPERATOR NAME:
MICHAEL S CARBONE
Permit #: 50609
Operator Agency/Dept: RAMAPO TOWN PD
Reference Standard #: 23350

Nominal = 0.10 g/210L

BLANK TEST	0.00	00:39
INTERNAL STANDARD	VERIFIED	00:39
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	00:39
BLANK TEST	0.00	00:40
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	00:40
BLANK TEST	0.00	00:41
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	00:42
BLANK TEST	0.00	00:42

1/11/24
MSC



SUPERVISOR TEST

RAMAPO TOWN
POLICE DEPT
DATAMASTER dmt: 116306

Date: 01/18/2024




OPERATOR NAME:
MICHAEL S CARBONE
Permit #: 50609
Operator Agency/Dept: RAMAPO TOWN PD
Reference Standard #: 23350

Nominal = 0.10 g/210L

BLANK TEST	0.00	02:53
INTERNAL STANDARD	VERIFIED	02:53
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	02:53
BLANK TEST	0.00	02:54
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	02:54
BLANK TEST	0.00	02:55
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	02:55
BLANK TEST	0.00	02:56

1/18/24

MC520



SUPERVISOR TEST

RAMAPO TOWN
POLICE DEPT
DATAMASTER dmt: 116306




Date: 01/24/2024

OPERATOR NAME:
MICHAEL S CARBONE
Permit #: 50609
Operator Agency/Dept: RAMAPO TOWN PD
Reference Standard #: 23350

Nominal = 0.10 g/210L

BLANK TEST	0.00	23:46
INTERNAL STANDARD	VERIFIED	23:46
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	23:47
BLANK TEST	0.00	23:47
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	23:48
BLANK TEST	0.00	23:48
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	23:49
BLANK TEST	0.00	23:50

1/24/24



SOLUTION CHANGE

RAMAPO TOWN
POLICE DEPT
DATAMASTER dmt: 116306

Date: 02/02/2024




OPERATOR NAME:
MICHAEL S CARBONE
Permit #: 50609
Operator Agency/Dept: RAMAPO TOWN PD
Reference Standard #: 23350

Nominal = 0.10 g/210L

BLANK TEST	0.00	02:38
INTERNAL STANDARD	VERIFIED	02:38
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	02:38
BLANK TEST	0.00	02:39
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	02:39
BLANK TEST	0.00	02:40
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	02:40
BLANK TEST	0.00	02:41

2/2/24

MS20



SUPERVISOR TEST

RAMAPO TOWN
POLICE DEPT
DATAMASTER dmt: 116306




Date: 02/10/2024

OPERATOR NAME:
MICHAEL S CARBONE
Permit #: 50609
Operator Agency/Dept: RAMAPO TOWN PD
Reference Standard #: 23350

Nominal = 0.10 g/210L

BLANK TEST	0.00	00:23
INTERNAL STANDARD	VERIFIED	00:24
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	00:24
BLANK TEST	0.00	00:25
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	00:25
BLANK TEST	0.00	00:26
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	00:26
BLANK TEST	0.00	00:27

2/10/24
MC520



SUPERVISOR TEST

RAMAPO TOWN
POLICE DEPT
DATAMASTER dmt: 116306




Date: 02/17/2024

OPERATOR NAME:
MICHAEL S CARBONE
Permit #: 50609
Operator Agency/Dept: RAMAPO TOWN PD
Reference Standard #: 23350

Nominal = 0.10 g/210L

BLANK TEST	0.00	23:29
INTERNAL STANDARD	VERIFIED	23:29
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	23:30
BLANK TEST	0.00	23:30
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	23:31
BLANK TEST	0.00	23:32
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	23:32
BLANK TEST	0.00	23:33

MC520
2/17/24



SUPERVISOR TEST

RAMAPO TOWN
POLICE DEPT
DATAMASTER dmt: 116306

Date: 02/25/2024


OPERATOR NAME:
MICHAEL S CARBONE
Permit #: 50609
Operator Agency/Dept: RAMAPO TOWN PD
Reference Standard #: 23350

Nominal = 0.10 g/210L

BLANK TEST	0.00	07:53
INTERNAL STANDARD	VERIFIED	07:53
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	07:53
BLANK TEST	0.00	07:54
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	07:54
BLANK TEST	0.00	07:55
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	07:56
BLANK TEST	0.00	07:56

2/25/24

MSC



SOLUTION CHANGE

RAMAPO TOWN
 POLICE DEPT
 DATAMASTER dmt: 116306




Date: 03/03/2024

OPERATOR NAME:
 MICHAEL S CARBONE
 Permit #: 50609
 Operator Agency/Dept: RAMAPO TOWN PD
 Reference Standard #: 23430

Nominal = 0.10 g/210L

BLANK TEST	0.00	03:39
INTERNAL STANDARD	VERIFIED	03:39
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	03:40
BLANK TEST	0.00	03:40
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	03:41
BLANK TEST	0.00	03:42
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	03:42
BLANK TEST	0.00	03:43

3/3/24
 MCB



SUPERVISOR TEST

RAMAPO TOWN
 POLICE DEPT
 DATAMASTER dmt: 116306


Date: 03/12/2024

OPERATOR NAME:
 MICHAEL S CARBONE
 Permit #: 50609
 Operator Agency/Dept: RAMAPO TOWN PD
 Reference Standard #: 23430

Nominal = 0.10 g/210L

BLANK TEST	0.00	02:05
INTERNAL STANDARD	VERIFIED	02:05
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	02:05
BLANK TEST	0.00	02:06
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	02:06
BLANK TEST	0.00	02:07
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	02:07
BLANK TEST	0.00	02:08

3/12/24
McCaro



SUPERVISOR TEST

RAMAPO TOWN
POLICE DEPT
DATAMASTER dmt: 116306


Date: 03/18/2024

OPERATOR NAME:
MICHAEL S CARBONE
Permit #: 50609
Operator Agency/Dept: RAMAPO TOWN PD
Reference Standard #: 23430

Nominal = 0.10 g/210L

BLANK TEST	0.00	01:41
INTERNAL STANDARD	VERIFIED	01:41
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	01:42
BLANK TEST	0.00	01:42
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	01:43
BLANK TEST	0.00	01:43
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	01:44
BLANK TEST	0.00	01:45

3/18/24
MC



SUPERVISOR TEST

RAMAPO TOWN
 POLICE DEPT
 DATAMASTER dmt: 116306

Date: 03/27/2024


OPERATOR NAME:
 MICHAEL S CARBONE
 Permit #: 50609
 Operator Agency/Dept: RAMAPO TOWN PD
 Reference Standard #: 23430

Nominal = 0.10 g/210L

BLANK TEST	0.00	01:20
INTERNAL STANDARD	VERIFIED	01:20
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	01:20
BLANK TEST	0.00	01:21
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	01:21
BLANK TEST	0.00	01:22
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	01:22
BLANK TEST	0.00	01:23

3/27/24

MS 20



SOLUTION CHANGE

RAMAPO TOWN
POLICE DEPT
DATAMASTER dmt: 116306

Date: 04/05/2024




OPERATOR NAME:
MICHAEL S CARBONE
Permit #: 50609
Operator Agency/Dept: RAMAPO TOWN PD
Reference Standard #: 23430

Nominal = 0.10 g/210L

BLANK TEST	0.00	04:49
INTERNAL STANDARD	VERIFIED	04:50
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	04:50
BLANK TEST	0.00	04:51
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	04:51
BLANK TEST	0.00	04:52
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	04:52
BLANK TEST	0.00	04:53

4/5/24

MSC



SUPERVISOR TEST

RAMAPO TOWN
POLICE DEPT
DATAMASTER dmt: 116306

Date: 04/13/2024




OPERATOR NAME:
MICHAEL S CARBONE
Permit #: 50609
Operator Agency/Dept: RAMAPO TOWN PD
Reference Standard #: 23430

Nominal = 0.10 g/210L

BLANK TEST	0.00	05:44
INTERNAL STANDARD	VERIFIED	05:44
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	05:44
BLANK TEST	0.00	05:45
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	05:45
BLANK TEST	0.00	05:46
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	05:46
BLANK TEST	0.00	05:47

4/13/24

Mc 520



SUPERVISOR TEST

RAMAPO TOWN
POLICE DEPT
DATAMASTER dmt: 116306

Date: 04/24/2024




OPERATOR NAME:
MICHAEL S CARBONE
Permit #: 50609
Operator Agency/Dept: RAMAPO TOWN PD
Reference Standard #: 23430

Nominal = 0.10 g/210L

BLANK TEST	0.00	00:02
INTERNAL STANDARD	VERIFIED	00:02
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	00:02
BLANK TEST	0.00	00:03
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	00:03
BLANK TEST	0.00	00:04
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	00:04
BLANK TEST	0.00	00:05

4/24/24

MSC



SOLUTION CHANGE

RAMAPO TOWN
POLICE DEPT
DATAMASTER dmt: 116306




Date: 05/02/2024

OPERATOR NAME:
MICHAEL S CARBONE
Permit #: 50609
Operator Agency/Dept: RAMAPO TOWN PD
Reference Standard #: 24040

Nominal = 0.10 g/210L

BLANK TEST	0.00	00:05
INTERNAL STANDARD	VERIFIED	00:05
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	00:06
BLANK TEST	0.00	00:06
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	00:07
BLANK TEST	0.00	00:08
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	00:08
BLANK TEST	0.00	00:09

5/2/24
MC520



SUPERVISOR TEST

RAMAPO TOWN
POLICE DEPT
DATAMASTER dmt: 116306




Date: 05/09/2024

OPERATOR NAME:
MICHAEL S CARBONE
Permit #: 50609
Operator Agency/Dept: RAMAPO TOWN PD
Reference Standard #: 24040

Nominal = 0.10 g/210L

BLANK TEST	0.00	06:55
INTERNAL STANDARD	VERIFIED	06:55
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	06:55
BLANK TEST	0.00	06:56
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	06:56
BLANK TEST	0.00	06:57
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	06:57
BLANK TEST	0.00	06:58

MC520
5/9/24



SUPERVISOR TEST

RAMAPO TOWN
POLICE DEPT
DATAMASTER dmt: 116306




Date: 05/18/2024

OPERATOR NAME:
MICHAEL S CARBONE
Permit #: 50609
Operator Agency/Dept: RAMAPO TOWN PD
Reference Standard #: 24040

Nominal = 0.10 g/210L

BLANK TEST	0.00	00:42
INTERNAL STANDARD	VERIFIED	00:42
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	00:42
BLANK TEST	0.00	00:43
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	00:43
BLANK TEST	0.00	00:44
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	00:45
BLANK TEST	0.00	00:45

5/18/24
MCS



SUPERVISOR TEST

RAMAPO TOWN
POLICE DEPT
DATAMASTER dmt: 116306




Date: 05/28/2024

OPERATOR NAME:
MICHAEL S CARBONE
Permit #: 50609
Operator Agency/Dept: RAMAPO TOWN PD
Reference Standard #: 24040

Nominal = 0.10 g/210L

BLANK TEST	0.00	01:54
INTERNAL STANDARD	VERIFIED	01:54
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	01:55
BLANK TEST	0.00	01:56
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	01:56
BLANK TEST	0.00	01:57
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	01:57
BLANK TEST	0.00	01:58

5/28/24
MC 520



SOLUTION CHANGE

RAMAPO TOWN
POLICE DEPT
DATAMASTER dmt: 116306




Date: 06/01/2024

OPERATOR NAME:
MICHAEL S CARBONE
Permit #: 50609
Operator Agency/Dept: RAMAPO TOWN PD
Reference Standard #: 24140

Nominal = 0.10 g/210L

BLANK TEST	0.00	01:35
INTERNAL STANDARD	VERIFIED	01:35
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	01:35
BLANK TEST	0.00	01:36
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	01:37
BLANK TEST	0.00	01:37
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	01:38
BLANK TEST	0.00	01:38

6/1/24
MC 520



SUPERVISOR TEST

RAMAPO TOWN
POLICE DEPT
DATAMASTER dmt: 116306




Date: 06/10/2024

OPERATOR NAME:
MICHAEL S CARBONE
Permit #: 50609
Operator Agency/Dept: RAMAPO TOWN PD
Reference Standard #: 24130

Nominal = 0.10 g/210L

BLANK TEST	0.00	06:37
INTERNAL STANDARD	VERIFIED	06:37
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	06:37
BLANK TEST	0.00	06:38
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	06:39
BLANK TEST	0.00	06:39
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	06:40
BLANK TEST	0.00	06:41

6/10/24
MS



SUPERVISOR TEST




RAMAPO TOWN
 POLICE DEPT
 DATAMASTER dmt: 116306

Date: 06/12/2024

OPERATOR NAME:
 MICHAEL S CARBONE
 Permit #: 50609
 Operator Agency/Dept: RAMAPO TOWN PD
 Reference Standard #: 24140

Nominal = 0.10 g/210L

BLANK TEST	0.00	03:55
INTERNAL STANDARD	VERIFIED	03:55
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	03:55
BLANK TEST	0.00	03:56
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	03:56
BLANK TEST	0.00	03:57
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	03:57
BLANK TEST	0.00	03:58



SUPERVISOR TEST

RAMAPO TOWN
POLICE DEPT
DATAMASTER dmt: 116306




Date: 06/16/2024

OPERATOR NAME:
MICHAEL CARBONE
Permit #: 50609
Operator Agency/Dept: RAMAPO TOWN PD
Reference Standard #: 24140

Nominal = 0.10 g/210L

BLANK TEST	0.00	14:15
INTERNAL STANDARD	VERIFIED	14:15
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	14:16
BLANK TEST	0.00	14:16
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	14:17
BLANK TEST	0.00	14:18
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	14:18
BLANK TEST	0.00	14:19

6/16/24
MC520



SUPERVISOR TEST

RAMAPO TOWN
POLICE DEPT
DATAMASTER dmt: 116306

Date: 06/26/2024




OPERATOR NAME:
MICHAEL S CARBONE
Permit #: 50609
Operator Agency/Dept: RAMAPO TOWN PD
Reference Standard #: 24140

Nominal = 0.10 g/210L

BLANK TEST	0.00	00:12
INTERNAL STANDARD	VERIFIED	00:12
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	00:12
BLANK TEST	0.00	00:13
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	00:13
BLANK TEST	0.00	00:14
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	00:14
BLANK TEST	0.00	00:15

6/26/24

McCrone



SOLUTION CHANGE

RAMAPO TOWN
POLICE DEPT
DATAMASTER dmt: 116306




Date: 06/28/2024

OPERATOR NAME:
MICHAEL S CARBONE
Permit #: 50609
Operator Agency/Dept: RAMAPO TOWN PD
Reference Standard #: 24140

Nominal = 0.10 g/210L

BLANK TEST	0.00	09:13
INTERNAL STANDARD	VERIFIED	09:13
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	09:13
BLANK TEST	0.00	09:14
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	09:15
BLANK TEST	0.00	09:15
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	09:16
BLANK TEST	0.00	09:17

6/28/24
MC 520



SUPERVISOR TEST

RAMAPO TOWN
POLICE DEPT
DATAMASTER dmt: 116306

Date: 07/05/2024




OPERATOR NAME:
MICHAEL S CARBONE
Permit #: 50609
Operator Agency/Dept: RAMAPO TOWN PD
Reference Standard #: 24140

Nominal = 0.10 g/210L

BLANK TEST	0.00	01:24
INTERNAL STANDARD	VERIFIED	01:24
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	01:24
BLANK TEST	0.00	01:25
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	01:25
BLANK TEST	0.00	01:26
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	01:26
BLANK TEST	0.00	01:27

7/5/24

MS20



SUPERVISOR TEST

RAMAPO TOWN
POLICE DEPT
DATAMASTER dmt: 116306




Date: 07/13/2024

OPERATOR NAME:
MICHAEL S CARBONE
Permit #: 50609
Operator Agency/Dept: RAMAPO TOWN PD
Reference Standard #: 24140

Nominal = 0.10 g/210L

BLANK TEST	0.00	23:18
INTERNAL STANDARD	VERIFIED	23:18
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	23:19
BLANK TEST	0.00	23:19
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	23:20
BLANK TEST	0.00	23:20
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	23:21
BLANK TEST	0.00	23:22

7/13/24
MC 520



SUPERVISOR TEST

RAMAPO TOWN
POLICE DEPT
DATAMASTER dmt: 116306




Date: 07/23/2024

OPERATOR NAME:
MICHAEL S CARBONE
Permit #: 50609
Operator Agency/Dept: RAMAPO TOWN PD
Reference Standard #: 24140

Nominal = 0.10 g/210L

BLANK TEST	0.00	00:39
INTERNAL STANDARD	VERIFIED	00:39
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	00:39
BLANK TEST	0.00	00:40
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	00:40
BLANK TEST	0.00	00:41
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	00:42
BLANK TEST	0.00	00:42

7/23/24
MCS20



SOLUTION CHANGE

RAMAPO TOWN
POLICE DEPT
DATAMASTER dmt: 116306




Date: 08/01/2024

OPERATOR NAME:
MICHAEL S CARBONE
Permit #: 50609
Operator Agency/Dept: RAMAPO TOWN PD
Reference Standard #: 24190

Nominal = 0.10 g/210L

BLANK TEST	0.00	00:31
INTERNAL STANDARD	VERIFIED	00:31
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	00:31
BLANK TEST	0.00	00:32
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	00:33
BLANK TEST	0.00	00:33
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	00:34
BLANK TEST	0.00	00:35

8/1/24
MCS20



SUPERVISOR TEST

RAMAPO TOWN
POLICE DEPT
DATAMASTER dmt: 116306

Date: 08/08/2024

OPERATOR NAME:
MICHAEL S CARBONE
Permit #: 50609
Operator Agency/Dept: RAMAPO TOWN PD
Reference Standard #: 24190

Nominal = 0.10 g/210L

BLANK TEST	0.00	00:10
INTERNAL STANDARD	VERIFIED	00:10
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	00:11
BLANK TEST	0.00	00:12
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	00:12
BLANK TEST	0.00	00:13
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	00:13
BLANK TEST	0.00	00:14

8/8/24
MCS20

